

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP031125

Review By	yogesh	Review On	3/12/2025 7:57:33 AM
Supervise By	mohammad	Supervise On	3/19/2025 6:54:29 AM
SubDirectory	PP031125	HP Acquire Method	HP Processing Method PP031125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP070417.D	11 Mar 2025 14:41		YP\AJ	Ok
2	I.BLK	I.BLK	PP070418.D	11 Mar 2025 14:57		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP070419.D	11 Mar 2025 15:13		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP070420.D	11 Mar 2025 15:29		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP070421.D	11 Mar 2025 15:46		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP070422.D	11 Mar 2025 16:02		YP\AJ	Ok,M
7	AR1660ICC050	AR1660ICC050	PP070423.D	11 Mar 2025 16:18		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP070424.D	11 Mar 2025 16:34		YP\AJ	Ok
9	AR1232ICC1000	AR1232ICC1000	PP070425.D	11 Mar 2025 16:51		YP\AJ	Ok
10	AR1232ICC750	AR1232ICC750	PP070426.D	11 Mar 2025 17:07		YP\AJ	Ok
11	AR1232ICC500	AR1232ICC500	PP070427.D	11 Mar 2025 17:23		YP\AJ	Ok
12	AR1232ICC250	AR1232ICC250	PP070428.D	11 Mar 2025 17:40		YP\AJ	Ok,M
13	AR1232ICC050	AR1232ICC050	PP070429.D	11 Mar 2025 17:56		YP\AJ	Ok,M
14	AR1242ICC1000	AR1242ICC1000	PP070430.D	11 Mar 2025 18:12		YP\AJ	Ok
15	AR1242ICC750	AR1242ICC750	PP070431.D	11 Mar 2025 18:28		YP\AJ	Ok
16	AR1242ICC500	AR1242ICC500	PP070432.D	11 Mar 2025 18:45		YP\AJ	Ok
17	AR1242ICC250	AR1242ICC250	PP070433.D	11 Mar 2025 19:01		YP\AJ	Ok
18	AR1242ICC050	AR1242ICC050	PP070434.D	11 Mar 2025 19:17		YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP031125

Review By	yogesh	Review On	3/12/2025 7:57:33 AM
Supervise By	mohammad	Supervise On	3/19/2025 6:54:29 AM
SubDirectory	PP031125	HP Acquire Method	HP Processing Method PP031125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC1000	AR1248ICC1000	PP070435.D	11 Mar 2025 19:34		YPIAJ	Ok
20	AR1248ICC750	AR1248ICC750	PP070436.D	11 Mar 2025 19:50		YPIAJ	Ok
21	AR1248ICC500	AR1248ICC500	PP070437.D	11 Mar 2025 20:06		YPIAJ	Ok
22	AR1248ICC250	AR1248ICC250	PP070438.D	11 Mar 2025 20:22		YPIAJ	Ok
23	AR1248ICC050	AR1248ICC050	PP070439.D	11 Mar 2025 20:39		YPIAJ	Ok,M
24	AR1254ICC1000	AR1254ICC1000	PP070440.D	11 Mar 2025 20:55		YPIAJ	Ok
25	AR1254ICC750	AR1254ICC750	PP070441.D	11 Mar 2025 21:11		YPIAJ	Ok
26	AR1254ICC500	AR1254ICC500	PP070442.D	11 Mar 2025 21:28		YPIAJ	Ok
27	AR1254ICC250	AR1254ICC250	PP070443.D	11 Mar 2025 21:44		YPIAJ	Ok
28	AR1254ICC050	AR1254ICC050	PP070444.D	11 Mar 2025 22:00		YPIAJ	Ok
29	AR1262ICC500	AR1262ICC500	PP070445.D	11 Mar 2025 22:16		YPIAJ	Ok
30	AR1268ICC1000	AR1268ICC1000	PP070446.D	11 Mar 2025 22:33		YPIAJ	Ok
31	AR1268ICC750	AR1268ICC750	PP070447.D	11 Mar 2025 22:49		YPIAJ	Ok
32	AR1268ICC500	AR1268ICC500	PP070448.D	11 Mar 2025 23:05		YPIAJ	Ok
33	AR1268ICC250	AR1268ICC250	PP070449.D	11 Mar 2025 23:22		YPIAJ	Ok
34	AR1268ICC050	AR1268ICC050	PP070450.D	11 Mar 2025 23:38		YPIAJ	Ok,M
35	PP031125ICV500	ICVPP031125	PP070451.D	11 Mar 2025 23:54		YPIAJ	Ok
36	AR1232ICV500	ICVPP031125AR1232	PP070452.D	12 Mar 2025 00:10		YPIAJ	Ok
37	AR1242ICV500	ICVPP031125AR1242	PP070453.D	12 Mar 2025 00:27		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP031125

Review By	yogesh	Review On	3/12/2025 7:57:33 AM
Supervise By	mohammad	Supervise On	3/19/2025 6:54:29 AM
SubDirectory	PP031125	HP Acquire Method	HP Processing Method PP031125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	AR1248ICV500	ICVPP031125AR1248	PP070454.D	12 Mar 2025 00:43		YPIAJ	Ok
39	AR1254ICV500	ICVPP031125AR1254	PP070455.D	12 Mar 2025 00:59		YPIAJ	Ok
40	AR1268ICV500	ICVPP031125AR1268	PP070456.D	12 Mar 2025 01:16		YPIAJ	Ok
41	AR1660CCC500	AR1660CCC500	PP070457.D	12 Mar 2025 01:32		YPIAJ	Ok
42	AR1232CCC500	AR1232CCC500	PP070458.D	12 Mar 2025 01:48		YPIAJ	Ok
43	AR1242CCC500	AR1242CCC500	PP070459.D	12 Mar 2025 02:04		YPIAJ	Ok
44	I.BLK	I.BLK	PP070460.D	12 Mar 2025 02:21		YPIAJ	Ok
45	PB167066BL	PB167066BL	PP070461.D	12 Mar 2025 02:37		YPIAJ	Ok
46	Q1502-15	PT-PCBW-WP	PP070462.D	12 Mar 2025 02:53	AR1232 HIT	YPIAJ	Ok
47	AR1660CCC500	AR1660CCC500	PP070463.D	12 Mar 2025 03:09		YPIAJ	Ok
48	AR1232CCC500	AR1232CCC500	PP070464.D	12 Mar 2025 03:26		YPIAJ	Ok
49	AR1242CCC500	AR1242CCC500	PP070465.D	12 Mar 2025 03:42		YPIAJ	Ok,M
50	I.BLK	I.BLK	PP070466.D	12 Mar 2025 03:58		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP031925

Review By	yogesh	Review On	3/19/2025 11:52:51 AM
Supervise By		Supervise On	
SubDirectory	PP031925	HP Acquire Method	HP Processing Method PP031125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP070682.D	19 Mar 2025 08:39		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP070683.D	19 Mar 2025 08:55		YP\AJ	Ok
3	AR1242CCC500	AR1242CCC500	PP070684.D	19 Mar 2025 09:17		YP\AJ	Ok
4	AR1248CCC500	AR1248CCC500	PP070685.D	19 Mar 2025 09:34		YP\AJ	Ok
5	AR1254CCC500	AR1254CCC500	PP070686.D	19 Mar 2025 09:50		YP\AJ	Ok
6	I.BLK	I.BLK	PP070687.D	19 Mar 2025 10:06		YP\AJ	Ok
7	PB167201BL	PB167201BL	PP070688.D	19 Mar 2025 12:05		YP\AJ	Ok
8	PB167201BS	PB167201BS	PP070689.D	19 Mar 2025 12:21		YP\AJ	Ok
9	Q1599-01	1A-1B-1C	PP070690.D	19 Mar 2025 12:38		YP\AJ	Ok
10	Q1599-02	2A-2B-2C	PP070691.D	19 Mar 2025 12:54		YP\AJ	Ok,NS
11	Q1599-03	3A-3B-3C	PP070692.D	19 Mar 2025 13:10		YP\AJ	Ok
12	Q1599-04	4A-4B-4C	PP070693.D	19 Mar 2025 13:26		YP\AJ	Ok
13	Q1599-05	5A-5B-5C	PP070694.D	19 Mar 2025 13:43		YP\AJ	Ok
14	Q1599-06	6A-6B-6C	PP070695.D	19 Mar 2025 13:59		YP\AJ	Ok
15	Q1598-03	72-11944	PP070696.D	19 Mar 2025 14:15		YP\AJ	Ok
16	Q1598-01	ETGI-354	PP070697.D	19 Mar 2025 14:32	TCMX high in 1st column, AR1260 Hit	YP\AJ	Ok,NS
17	AR1660CCC500	AR1660CCC500	PP070698.D	19 Mar 2025 15:20		YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP031925

Review By	yogesh	Review On	3/19/2025 11:52:51 AM
Supervise By		Supervise On	
SubDirectory	PP031925	HP Acquire Method	HP Processing Method PP031125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1242CCC500	AR1242CCC500	PP070699.D	19 Mar 2025 15:37		YPIAJ	Ok
19	AR1248CCC500	AR1248CCC500	PP070700.D	19 Mar 2025 15:53		YPIAJ	Ok
20	AR1254CCC500	AR1254CCC500	PP070701.D	19 Mar 2025 16:09		YPIAJ	Ok
21	I.BLK	I.BLK	PP070702.D	19 Mar 2025 16:26		YPIAJ	Ok
22	PB167202BL	PB167202BL	PP070703.D	19 Mar 2025 16:42		YPIAJ	Ok
23	PB167202BS	PB167202BS	PP070704.D	19 Mar 2025 16:58		YPIAJ	Ok
24	Q1596-01	15D	PP070705.D	19 Mar 2025 17:14	AR1254 Hit, Need 2X Dilution	YPIAJ	Dilution
25	Q1596-01MS	15DMS	PP070706.D	19 Mar 2025 17:31	Recovery Fail for AR1260	YPIAJ	Ok,NS
26	Q1596-01MSD	15DMSD	PP070707.D	19 Mar 2025 17:47	Recovery Fail for AR1260	YPIAJ	Ok,NS
27	Q1596-02	16D	PP070708.D	19 Mar 2025 18:03	AR1254 Hit	YPIAJ	Ok,NS
28	Q1597-01	1-CONCRETE-SLAB	PP070709.D	19 Mar 2025 18:19		YPIAJ	Ok
29	Q1597-03	2-CONCRETE-SLAB	PP070710.D	19 Mar 2025 18:36		YPIAJ	Ok
30	Q1597-05	3-CONCRETE-SLAB	PP070711.D	19 Mar 2025 18:52		YPIAJ	Ok,NS
31	AR1660CCC500	AR1660CCC500	PP070712.D	19 Mar 2025 19:41		YPIAJ	Ok
32	AR1242CCC500	AR1242CCC500	PP070713.D	19 Mar 2025 19:57		YPIAJ	Ok,NS
33	AR1248CCC500	AR1248CCC500	PP070714.D	19 Mar 2025 20:13		YPIAJ	Ok
34	AR1254CCC500	AR1254CCC500	PP070715.D	19 Mar 2025 20:29		YPIAJ	Ok
35	I.BLK	I.BLK	PP070716.D	19 Mar 2025 20:46		YPIAJ	Ok
36	PB167217BL	PB167217BL	PP070717.D	19 Mar 2025 21:02		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP031925

Review By	yogesh	Review On	3/19/2025 11:52:51 AM
Supervise By		Supervise On	
SubDirectory	PP031925	HP Acquire Method	HP Processing Method PP031125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

37	PB167217BS	PB167217BS	PP070718.D	19 Mar 2025 21:18		YPIAJ	Ok
38	Q1601-01	MP-1	PP070719.D	19 Mar 2025 21:34	AR1242 Hit, Need 2X dilution	YPIAJ	Dilution
39	Q1601-02	MP-2	PP070720.D	19 Mar 2025 21:51	AR1242 Hit, carry over, run again for confirmation	YPIAJ	Not Ok
40	Q1601-03	MP-3	PP070721.D	19 Mar 2025 22:07	AR1242 Hit,	YPIAJ	Ok,NS
41	Q1602-01	HOUSE-WIPE	PP070722.D	19 Mar 2025 22:23	AR1242 Hit, Need 4X Dilution	YPIAJ	Dilution
42	AR1660CCC500	AR1660CCC500	PP070723.D	19 Mar 2025 23:12		YPIAJ	Ok,NS
43	AR1242CCC500	AR1242CCC500	PP070724.D	19 Mar 2025 23:28		YPIAJ	Ok,NS
44	AR1248CCC500	AR1248CCC500	PP070725.D	19 Mar 2025 23:44		YPIAJ	Ok
45	AR1254CCC500	AR1254CCC500	PP070726.D	20 Mar 2025 00:00		YPIAJ	Ok,NS
46	I.BLK	I.BLK	PP070727.D	20 Mar 2025 00:17		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP032025

Review By	yogesh	Review On	3/20/2025 8:46:48 AM
Supervise By		Supervise On	
SubDirectory	PP032025	HP Acquire Method	HP Processing Method PP031125

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP070728.D	20 Mar 2025 07:11		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP070729.D	20 Mar 2025 07:27		YPIAJ	Ok,NS
3	AR1242CCC500	AR1242CCC500	PP070730.D	20 Mar 2025 07:45		YPIAJ	Ok,NS
4	AR1248CCC500	AR1248CCC500	PP070731.D	20 Mar 2025 08:02		YPIAJ	Ok
5	AR1254CCC500	AR1254CCC500	PP070732.D	20 Mar 2025 08:18		YPIAJ	Ok,NS
6	I.BLK	I.BLK	PP070733.D	20 Mar 2025 08:34		YPIAJ	Ok
7	Q1601-02	MP-2	PP070734.D	20 Mar 2025 08:50	AR1242 Hit	YPIAJ	Ok,NS
8	Q1601-01DL	MP-1DL	PP070735.D	20 Mar 2025 09:07	AR1242 Hit	YPIAJ	Ok,NS
9	Q1602-01DL	HOUSE-WIPEDL	PP070736.D	20 Mar 2025 09:23	AR1242 Hit	YPIAJ	Ok,NS
10	Q1596-01DL	15DDL	PP070737.D	20 Mar 2025 09:39	AR1254 Hit	YPIAJ	Ok,NS
11	Q1602-02	HOUSE-PAINT	PP070738.D	20 Mar 2025 11:53	AR1254 Hit & Need 10X, DCB high 2nd col	YPIAJ	Dilution
12	Q1605-01	DRUM-SOIL-CUTTING	PP070739.D	20 Mar 2025 12:09		YPIAJ	Ok
13	Q1606-01	CHRT24743	PP070740.D	20 Mar 2025 12:25		YPIAJ	Ok,NS
14	Q1606-03	RBR251346	PP070741.D	20 Mar 2025 12:42	AR1242 Hit	YPIAJ	Ok,NS
15	Q1606-05	RT4534	PP070742.D	20 Mar 2025 12:58	AR1254 Hit	YPIAJ	Ok,NS
16	Q1602-02DL	HOUSE-PAINTDL	PP070743.D	20 Mar 2025 13:14	DCB High in 2nd column, AR1254 Hit	YPIAJ	Ok,NS
17	AR1660CCC500	AR1660CCC500	PP070744.D	20 Mar 2025 14:03		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP032025

Review By	yogesh	Review On	3/20/2025 8:46:48 AM
Supervise By		Supervise On	
SubDirectory	PP032025	HP Acquire Method	HP Processing Method PP031125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1242CCC500	AR1242CCC500	PP070745.D	20 Mar 2025 14:19		YPIAJ	Ok,NS
19	AR1248CCC500	AR1248CCC500	PP070746.D	20 Mar 2025 14:35		YPIAJ	Ok,NS
20	AR1254CCC500	AR1254CCC500	PP070747.D	20 Mar 2025 14:52		YPIAJ	Ok,NS
21	I.BLK	I.BLK	PP070748.D	20 Mar 2025 15:08		YPIAJ	Ok
22	PB167222BS	PB167222BS	PP070750.D	20 Mar 2025 15:40		YPIAJ	Ok,NS
23	Q1603-01	NB-070-031925	PP070751.D	20 Mar 2025 15:57		YPIAJ	Ok,NS
24	Q1603-01MS	NB-070-031925MS	PP070752.D	20 Mar 2025 16:13		YPIAJ	Ok,NS
25	Q1603-01MSD	NB-070-031925MSD	PP070753.D	20 Mar 2025 16:29		YPIAJ	Ok,NS
26	Q1607-01	OR-03-031925	PP070754.D	20 Mar 2025 16:45		YPIAJ	Ok
27	Q1606-09	CHRT28607	PP070755.D	20 Mar 2025 17:02		YPIAJ	Ok,NS
28	Q1606-07	RT3025	PP070756.D	20 Mar 2025 17:18	AR1254 + AR1260 Hit	YPIAJ	Ok,NS
29	AR1660CCC500	AR1660CCC500	PP070757.D	20 Mar 2025 18:07		YPIAJ	Ok,NS
30	AR1242CCC500	AR1242CCC500	PP070758.D	20 Mar 2025 18:23		YPIAJ	Ok
31	AR1248CCC500	AR1248CCC500	PP070759.D	20 Mar 2025 18:39		YPIAJ	Ok
32	AR1254CCC500	AR1254CCC500	PP070760.D	20 Mar 2025 18:55		YPIAJ	Ok,NS
33	I.BLK	I.BLK	PP070761.D	20 Mar 2025 19:12		YPIAJ	Ok
34	PB167236BL	PB167236BL	PP070762.D	20 Mar 2025 19:28		YPIAJ	Ok
35	PB167236BS	PB167236BS	PP070763.D	20 Mar 2025 19:44		YPIAJ	Ok,NS
36	PB167236BSD	PB167236BSD	PP070764.D	20 Mar 2025 20:01		YPIAJ	Ok,NS

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP032025

Review By	yogesh	Review On	3/20/2025 8:46:48 AM
Supervise By		Supervise On	
SubDirectory	PP032025	HP Acquire Method	HP Processing Method PP031125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

37	Q1604-02	FRAC-TANK-N45878	PP070765.D	20 Mar 2025 20:17		YPIAJ	Ok
38	Q1606-11	N48965	PP070766.D	20 Mar 2025 20:33	AR1260 Hit, Need 5X Dilution	YPIAJ	Dilution
39	Q1616-01	7A-7B-7C	PP070767.D	20 Mar 2025 20:49		YPIAJ	Ok,NS
40	Q1617-01	1A-1B-1C 1957 BLG	PP070768.D	20 Mar 2025 21:06	AR1254 Hit	YPIAJ	Ok,NS
41	Q1612-01	BC260105-1-1	PP070769.D	20 Mar 2025 21:22	AR1254 Hit, DCB High in 1st column	YPIAJ	Ok,NS
42	Q1612-02	BC260105-1-2	PP070770.D	20 Mar 2025 21:38	AR1254 Hit	YPIAJ	Ok,NS
43	AR1660CCC500	AR1660CCC500	PP070771.D	20 Mar 2025 22:27		YPIAJ	Ok,NS
44	AR1242CCC500	AR1242CCC500	PP070772.D	20 Mar 2025 22:43		YPIAJ	Ok,NS
45	AR1248CCC500	AR1248CCC500	PP070773.D	20 Mar 2025 23:00		YPIAJ	Ok
46	AR1254CCC500	AR1254CCC500	PP070774.D	20 Mar 2025 23:16		YPIAJ	Ok,NS
47	I.BLK	I.BLK	PP070775.D	20 Mar 2025 23:32		YPIAJ	Ok
48	PB167237BL	PB167237BL	PP070776.D	20 Mar 2025 23:49		YPIAJ	Ok
49	PB167237BS	PB167237BS	PP070777.D	21 Mar 2025 00:05		YPIAJ	Ok,NS
50	PB167238BL	PB167238BL	PP070778.D	21 Mar 2025 00:21		YPIAJ	Ok
51	PB167238BS	PB167238BS	PP070779.D	21 Mar 2025 00:38		YPIAJ	Ok,NS
52	PB167239BL	PB167239BL	PP070780.D	21 Mar 2025 00:54		YPIAJ	Ok
53	PB167239BS	PB167239BS	PP070781.D	21 Mar 2025 01:10		YPIAJ	Ok,NS
54	Q1612-03	BC216341-1-1	PP070782.D	21 Mar 2025 01:26	AR1254 Hit	YPIAJ	Ok,NS
55	Q1612-04	BC216341-1-2	PP070783.D	21 Mar 2025 01:43	AR1254 Hit	YPIAJ	Ok,NS

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP032025

Review By	yogesh	Review On	3/20/2025 8:46:48 AM
Supervise By		Supervise On	
SubDirectory	PP032025	HP Acquire Method	HP Processing Method PP031125
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

56	Q1613-01	LAW-25-0051	PP070784.D	21 Mar 2025 01:59	DCB High in 1st column	YP\AJ	Ok,NS
57	AR1660CCC500	AR1660CCC500	PP070785.D	21 Mar 2025 03:09		YP\AJ	Ok,NS
58	AR1242CCC500	AR1242CCC500	PP070786.D	21 Mar 2025 03:26		YP\AJ	Ok,NS
59	AR1248CCC500	AR1248CCC500	PP070787.D	21 Mar 2025 03:42		YP\AJ	Ok,NS
60	AR1254CCC500	AR1254CCC500	PP070788.D	21 Mar 2025 03:58	4th peak high in 1st column	YP\AJ	Ok,NS
61	I.BLK	I.BLK	PP070789.D	21 Mar 2025 04:15		YP\AJ	Ok

M : Manual Integration