

Surrogate Summary

SDG No.: **Q1596**

Client: **ATC Group Services LLC**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP070418.D	PIBLK-PP070418.D	Tetrachloro-m-xylene	1	20	22.0	110		60	140
		Decachlorobiphenyl	1	20	23.1	116		60	140
		Tetrachloro-m-xylene	2	20	23.0	115		60	140
		Decachlorobiphenyl	2	20	23.5	117		60	140
I.BLK-PP070702.D	PIBLK-PP070702.D	Tetrachloro-m-xylene	1	20	20.1	101		60	140
		Decachlorobiphenyl	1	20	21.8	109		60	140
		Tetrachloro-m-xylene	2	20	21.9	109		60	140
		Decachlorobiphenyl	2	20	20.9	104		60	140
PB167202BL	PB167202BL	Tetrachloro-m-xylene	1	20	21.1	105		32	144
		Decachlorobiphenyl	1	20	21.5	108		32	175
		Tetrachloro-m-xylene	2	20	22.3	112		32	144
		Decachlorobiphenyl	2	20	21.0	105		32	175
PB167202BS	PB167202BS	Tetrachloro-m-xylene	1	20	21.3	106		32	144
		Decachlorobiphenyl	1	20	22.2	111		32	175
		Tetrachloro-m-xylene	2	20	22.1	111		32	144
		Decachlorobiphenyl	2	20	20.3	101		32	175
Q1596-01	15D	Tetrachloro-m-xylene	1	20	20.6	103		32	144
		Decachlorobiphenyl	1	20	16.4	82		32	175
		Tetrachloro-m-xylene	2	20	20.7	104		32	144
		Decachlorobiphenyl	2	20	16.1	80		32	175
Q1596-01MS	15DMS	Tetrachloro-m-xylene	1	20	21.5	108		32	144
		Decachlorobiphenyl	1	20	16.4	82		32	175
		Tetrachloro-m-xylene	2	20	21.3	106		32	144
		Decachlorobiphenyl	2	20	17.7	88		32	175
Q1596-01MSD	15DMSD	Tetrachloro-m-xylene	1	20	21.3	107		32	144
		Decachlorobiphenyl	1	20	16.0	80		32	175
		Tetrachloro-m-xylene	2	20	20.7	104		32	144
		Decachlorobiphenyl	2	20	17.1	86		32	175
Q1596-02	16D	Tetrachloro-m-xylene	1	20	21.3	106		32	144
		Decachlorobiphenyl	1	20	14.5	72		32	175
		Tetrachloro-m-xylene	2	20	21.6	108		32	144
		Decachlorobiphenyl	2	20	15.4	77		32	175
I.BLK-PP070716.D	PIBLK-PP070716.D	Tetrachloro-m-xylene	1	20	20.6	103		60	140
		Decachlorobiphenyl	1	20	22.1	110		60	140
		Tetrachloro-m-xylene	2	20	22.9	115		60	140
		Decachlorobiphenyl	2	20	20.2	101		60	140
I.BLK-PP070733.D	PIBLK-PP070733.D	Tetrachloro-m-xylene	1	20	21.3	106		60	140
		Decachlorobiphenyl	1	20	22.9	114		60	140
		Tetrachloro-m-xylene	2	20	22.9	114		60	140
		Decachlorobiphenyl	2	20	21.1	106		60	140
Q1596-01DL	15DDL	Tetrachloro-m-xylene	1	20	19.8	99		32	144

Surrogate Summary

SDG No.: Q1596

Client: ATC Group Services LLC

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q1596-01DL	15DDL	Decachlorobiphenyl	1	20	16.9	85		32	175
		Tetrachloro-m-xylene	2	20	22.7	114		32	144
		Decachlorobiphenyl	2	20	16.6	83		32	175
I.BLK-PP070748.D	PIBLK-PP070748.D	Tetrachloro-m-xylene	1	20	20.7	103		60	140
		Decachlorobiphenyl	1	20	22.9	115		60	140
		Tetrachloro-m-xylene	2	20	23.5	117		60	140
		Decachlorobiphenyl	2	20	21.5	107		60	140