

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1599 SAS No.: Q1599 SDG NO.: Q1599

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL01 Date Analyzed: 03/19/2025

Lab Sample No.: AR1660CCC500 Data File : PP070683.D Time Analyzed: 08:55

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.670	5.577	5.777	458.300	500.000	-8.3
Aroclor-1016-2	5.691	5.599	5.799	519.580	500.000	3.9
Aroclor-1016-3	5.754	5.661	5.861	481.730	500.000	-3.7
Aroclor-1016-4	5.851	5.758	5.958	473.800	500.000	-5.2
Aroclor-1016-5	6.144	6.051	6.251	474.990	500.000	-5.0
Aroclor-1260-1	7.263	7.171	7.371	551.900	500.000	10.4
Aroclor-1260-2	7.516	7.425	7.625	544.580	500.000	8.9
Aroclor-1260-3	7.874	7.783	7.983	536.140	500.000	7.2
Aroclor-1260-4	8.098	8.008	8.208	508.880	500.000	1.8
Aroclor-1260-5	8.419	8.328	8.528	513.220	500.000	2.6
Decachlorobiphenyl	10.237	10.152	10.352	50.230	50.000	0.5
Tetrachloro-m-xylene	4.518	4.424	4.624	49.560	50.000	-0.9

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1599 SAS No.: Q1599 SDG NO.: Q1599

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL01 Date Analyzed: 03/19/2025

Lab Sample No.: AR1660CCC500 Data File : PP070683.D Time Analyzed: 08:55

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.909	4.817	5.017	505.040	500.000	1.0
Aroclor-1016-2	4.928	4.836	5.036	492.660	500.000	-1.5
Aroclor-1016-3	5.105	5.013	5.213	504.790	500.000	1.0
Aroclor-1016-4	5.146	5.054	5.254	517.130	500.000	3.4
Aroclor-1016-5	5.361	5.270	5.470	497.120	500.000	-0.6
Aroclor-1260-1	6.398	6.307	6.507	489.490	500.000	-2.1
Aroclor-1260-2	6.585	6.495	6.695	476.060	500.000	-4.8
Aroclor-1260-3	6.740	6.649	6.849	475.650	500.000	-4.9
Aroclor-1260-4	7.211	7.121	7.321	462.310	500.000	-7.5
Aroclor-1260-5	7.452	7.362	7.562	453.090	500.000	-9.4
Decachlorobiphenyl	8.867	8.781	8.981	46.440	50.000	-7.1
Tetrachloro-m-xylene	3.822	3.729	3.929	50.570	50.000	1.1