

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1599 SAS No.: Q1599 SDG NO.: Q1599

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/19/2025

Lab Sample No.: AR1660CCC500 Data File : PP070698.D Time Analyzed: 15:20

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.670	5.577	5.777	470.360	500.000	-5.9
Aroclor-1016-2	5.691	5.599	5.799	507.800	500.000	1.6
Aroclor-1016-3	5.753	5.661	5.861	481.900	500.000	-3.6
Aroclor-1016-4	5.851	5.758	5.958	476.500	500.000	-4.7
Aroclor-1016-5	6.144	6.051	6.251	465.840	500.000	-6.8
Aroclor-1260-1	7.262	7.171	7.371	540.410	500.000	8.1
Aroclor-1260-2	7.516	7.425	7.625	537.990	500.000	7.6
Aroclor-1260-3	7.874	7.783	7.983	530.050	500.000	6.0
Aroclor-1260-4	8.098	8.008	8.208	500.980	500.000	0.2
Aroclor-1260-5	8.419	8.328	8.528	511.080	500.000	2.2
Decachlorobiphenyl	10.235	10.152	10.352	49.600	50.000	-0.8
Tetrachloro-m-xylene	4.518	4.424	4.624	48.790	50.000	-2.4

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1599 SAS No.: Q1599 SDG NO.: Q1599

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/19/2025

Lab Sample No.: AR1660CCC500 Data File : PP070698.D Time Analyzed: 15:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.908	4.817	5.017	504.940	500.000	1.0
Aroclor-1016-2	4.927	4.836	5.036	502.660	500.000	0.5
Aroclor-1016-3	5.104	5.013	5.213	508.940	500.000	1.8
Aroclor-1016-4	5.145	5.054	5.254	520.950	500.000	4.2
Aroclor-1016-5	5.360	5.270	5.470	514.750	500.000	3.0
Aroclor-1260-1	6.397	6.307	6.507	486.030	500.000	-2.8
Aroclor-1260-2	6.585	6.495	6.695	475.080	500.000	-5.0
Aroclor-1260-3	6.738	6.649	6.849	471.530	500.000	-5.7
Aroclor-1260-4	7.210	7.121	7.321	452.330	500.000	-9.5
Aroclor-1260-5	7.451	7.362	7.562	452.300	500.000	-9.5
Decachlorobiphenyl	8.866	8.781	8.981	46.820	50.000	-6.4
Tetrachloro-m-xylene	3.822	3.729	3.929	51.730	50.000	3.5