

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1600

Project: Monthly 2025

Run No.: LB135111

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135111BS							
TPH	mg/L	20.0	16.8		84	1	78-114	03/20/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1600

Project: Monthly 2025

Run No.: LB135111

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135111BSD							
TPH	mg/L	20.0	17.0		85	1	78-114	03/20/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1600

Project: Monthly 2025

Run No.: LB135114

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135114BS							
BOD5	mg/L	198	185		94	1	84.6-115.4	03/20/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1600

Project: Monthly 2025

Run No.: LB135145

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135145BS							
TSS	mg/L	550	532		97	1	90-110	03/24/2025