

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1608

Client: AECOM Technical Services, Inc.

Analytical Method: 8270-Modified DataFile: BN036788.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB167251BSD	1,4-Dioxane	0.4	0.34	ug/L	85	0			70	130	20