

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q1608 SAS No.: Q1608 SDG No.: Q1608
 Instrument ID: MSVOA_X Calibration Date(s): 02/28/2025 02/28/2025
 Heated Purge: (Y/N) N Calibration Time(s): 01:27 03:47
 GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF001 = VX045068.D		RRF005 = VX045069.D		RRF020 = VX045070.D			
		RRF050 = VX045071.D		RRF100 = VX045072.D		RRF150 = VX045073.D			
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
Dichlorodifluoromethane	0.624	0.646	0.646	0.627	0.626	0.607	0.629	2.4	
Chloromethane	0.755	0.821	0.828	0.753	0.721	0.744	0.770	5.7	
Vinyl Chloride	0.773	0.755	0.774	0.761	0.765	0.758	0.764	1	
Bromomethane		0.337	0.298	0.292	0.284	0.291	0.300	7	
Chloroethane	0.373	0.421	0.366	0.373	0.297	0.286	0.352	14.5	
Trichlorofluoromethane	0.978	1.061	1.050	1.050	0.982	0.948	1.012	4.7	
1,1,2-Trichlorotrifluoroethane	0.526	0.613	0.595	0.594	0.596	0.563	0.581	5.4	
1,1-Dichloroethene	0.609	0.620	0.612	0.623	0.620	0.603	0.614	1.3	
Acetone	0.414	0.363	0.384	0.351	0.345	0.356	0.369	7	
Carbon Disulfide	1.584	1.582	1.587	1.660	1.708	1.698	1.636	3.6	
Methyl tert-butyl Ether	1.955	1.913	2.127	2.083	2.132	2.158	2.061	5	
Methyl Acetate	0.954	0.835	0.903	0.867	0.869	0.899	0.888	4.6	
Methylene Chloride	0.806	0.730	0.752	0.698	0.694	0.706	0.731	5.8	
trans-1,2-Dichloroethene	0.540	0.619	0.603	0.631	0.634	0.616	0.607	5.7	
1,1-Dichloroethane	1.200	1.223	1.280	1.242	1.270	1.264	1.247	2.5	
Cyclohexane		1.021	1.087	1.108	1.142	1.052	1.082	4.3	
2-Butanone	0.476	0.545	0.610	0.579	0.553	0.570	0.555	8.1	
Carbon Tetrachloride	0.463	0.463	0.447	0.468	0.489	0.463	0.465	3	
cis-1,2-Dichloroethene	0.687	0.746	0.765	0.762	0.767	0.769	0.749	4.2	
Bromochloromethane	0.557	0.633	0.613	0.603	0.563	0.596	0.594	5	
Chloroform	1.206	1.247	1.278	1.246	1.230	1.225	1.239	2	
1,1,1-Trichloroethane	0.908	0.992	1.009	1.024	1.044	1.025	1.000	4.8	
Methylcyclohexane	0.464	0.530	0.560	0.585	0.607	0.550	0.549	9.1	
Benzene	1.321	1.459	1.491	1.496	1.497	1.424	1.448	4.7	
1,2-Dichloroethane	0.487	0.525	0.545	0.528	0.524	0.520	0.521	3.6	
Trichloroethene	0.319	0.351	0.339	0.341	0.354	0.336	0.340	3.7	
1,2-Dichloropropane	0.354	0.378	0.382	0.371	0.376	0.373	0.372	2.7	
Bromodichloromethane	0.478	0.503	0.536	0.524	0.528	0.528	0.516	4.2	
4-Methyl-2-Pentanone	0.535	0.570	0.647	0.610	0.579	0.579	0.587	6.5	
Toluene	0.716	0.872	0.892	0.898	0.874	0.845	0.849	8	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q1608 SAS No.: Q1608 SDG No.: Q1608
 Instrument ID: MSVOA_X Calibration Date(s): 02/28/2025 02/28/2025
 Heated Purge: (Y/N) N Calibration Time(s): 01:27 03:47
 GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF001 = VX045068.D		RRF005 = VX045069.D		RRF020 = VX045070.D			
		RRF050 = VX045071.D		RRF100 = VX045072.D		RRF150 = VX045073.D			
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
t-1,3-Dichloropropene	0.304	0.389	0.436	0.469	0.490	0.502	0.431	17.3	
cis-1,3-Dichloropropene	0.404	0.463	0.509	0.535	0.555	0.553	0.503	11.8	
1,1,2-Trichloroethane	0.346	0.348	0.371	0.356	0.341	0.336	0.350	3.6	
2-Hexanone	0.349	0.412	0.476	0.448	0.431	0.436	0.425	10.1	
Dibromochloromethane	0.305	0.349	0.390	0.384	0.385	0.380	0.366	9	
1,2-Dibromoethane	0.311	0.352	0.371	0.356	0.355	0.350	0.349	5.7	
Tetrachloroethene	0.315	0.326	0.319	0.324	0.329	0.309	0.320	2.3	
Chlorobenzene	0.968	1.054	1.090	1.092	1.100	1.045	1.058	4.7	
Ethyl Benzene	1.566	1.794	1.889	1.952	1.972	1.888	1.843	8.1	
m/p-Xylenes	0.555	0.672	0.711	0.724	0.715	0.673	0.675	9.3	
o-Xylene	0.609	0.689	0.702	0.706	0.707	0.670	0.681	5.5	
Styrene	0.879	1.060	1.170	1.181	1.183	1.134	1.101	10.7	
Bromoform	0.209	0.234	0.276	0.276	0.300	0.300	0.266	13.9	
Isopropylbenzene	3.397	4.034	3.999	4.135	4.006	3.845	3.903	6.8	
1,1,2,2-Tetrachloroethane	1.395	1.479	1.513	1.419	1.391	1.396	1.432	3.6	
1,3-Dichlorobenzene	1.502	1.652	1.710	1.668	1.675	1.649	1.643	4.4	
1,4-Dichlorobenzene	1.605	1.702	1.665	1.687	1.669	1.643	1.662	2.1	
1,2-Dichlorobenzene	1.479	1.695	1.735	1.668	1.687	1.622	1.648	5.5	
1,2-Dibromo-3-Chloropropane	0.237	0.243	0.285	0.289	0.300	0.320	0.279	11.6	
1,2,4-Trichlorobenzene	0.668	0.821	0.978	1.009	1.074	1.080	0.938	17.3	
1,2,3-Trichlorobenzene	0.747	0.897	1.032	1.059	1.107	1.096	0.990	14.2	
1,2-Dichloroethane-d4		0.836	0.784	0.757	0.783	0.817	0.795	3.9	
Dibromofluoromethane		0.329	0.335	0.329	0.340	0.338	0.334	1.5	
Toluene-d8		1.237	1.191	1.210	1.219	1.203	1.212	1.4	
4-Bromofluorobenzene		0.383	0.393	0.402	0.410	0.421	0.402	3.7	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.