

Surrogate Summary

SDG No.: **Q1608**

Client: **AECOM Technical Services, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1608-29	TB	Toluene-d8	50	50.0	100	89	112
		4-Bromofluorobenzene	50	51.5	103	85	114
VX0321WBL01	VX0321WBL01	1,2-Dichloroethane-d4	50	56.9	114	81	118
		Dibromofluoromethane	50	53.4	107	80	119
		Toluene-d8	50	51.9	104	89	112
		4-Bromofluorobenzene	50	56.1	112	85	114
VX0321WBS01	VX0321WBS01	1,2-Dichloroethane-d4	50	54.8	110	81	118
		Dibromofluoromethane	50	54.1	108	80	119
		Toluene-d8	50	51.3	102	89	112
		4-Bromofluorobenzene	50	54.7	109	85	114
VX0324WBL01	VX0324WBL01	1,2-Dichloroethane-d4	50	58.0	116	81	118
		Dibromofluoromethane	50	53.6	107	80	119
		Toluene-d8	50	50.6	101	89	112
		4-Bromofluorobenzene	50	52.6	105	85	114
VX0324WBS01	VX0324WBS01	1,2-Dichloroethane-d4	50	57.0	114	81	118
		Dibromofluoromethane	50	56.2	112	80	119
		Toluene-d8	50	52.5	105	89	112
		4-Bromofluorobenzene	50	55.9	112	85	114
VX0325WBL01	VX0325WBL01	1,2-Dichloroethane-d4	50	58.3	117	81	118
		Dibromofluoromethane	50	54.2	108	80	119
		Toluene-d8	50	51.0	102	89	112
		4-Bromofluorobenzene	50	54.8	110	85	114
VX0325WBS02	VX0325WBS02	1,2-Dichloroethane-d4	50	56.6	113	81	118
		Dibromofluoromethane	50	53.4	107	80	119
		Toluene-d8	50	50.2	100	89	112
		4-Bromofluorobenzene	50	54.3	109	85	114