

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1608

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Parameter	Spike	Sample Result	Result	Units	Rec		RPD		Limits		RPD
					Rec	Qual	RPD	Qual	Low	High	
Lab Sample ID :	Q1608-12MS	Client Sample ID :	RE132D1-20250318MS						Datafile :	VX045453.D	
Dichlorodifluoromethane	50	0	63.1	ug/L	126				32	152	
Chloromethane	50	0	54.1	ug/L	108				50	139	
Vinyl chloride	50	0	49.4	ug/L	99				58	137	
Bromomethane	50	0	56.7	ug/L	113				53	141	
Chloroethane	50	0	56.0	ug/L	112				60	138	
Trichlorofluoromethane	50	0	53.8	ug/L	108				65	141	
1,1,2-Trichlorotrifluoroethane	50	4.90	62.5	ug/L	115				70	136	
1,1-Dichloroethene	50	0.74	53.5	ug/L	106				71	131	
Acetone	250	1.70	270	ug/L	107				39	160	
Carbon disulfide	50	0	47.8	ug/L	96				64	133	
Methyl tert-butyl Ether	50	0.76	53.4	ug/L	105				71	124	
Methyl Acetate	50	0	46.2	ug/L	92				56	136	
Methylene Chloride	50	0	50.6	ug/L	101				74	124	
trans-1,2-Dichloroethene	50	0	53.4	ug/L	107				75	124	
1,1-Dichloroethane	50	0	53.7	ug/L	107				77	125	
Cyclohexane	50	0	54.8	ug/L	110				71	130	
2-Butanone	250	0	280	ug/L	112				56	143	
Carbon Tetrachloride	50	0	55.6	ug/L	111				72	136	
cis-1,2-Dichloroethene	50	2.20	54.6	ug/L	105				78	123	
Bromochloromethane	50	0	55.1	ug/L	110				78	123	
Chloroform	50	2.10	56.5	ug/L	109				79	124	
1,1,1-Trichloroethane	50	0	56.4	ug/L	113				74	131	
Methylcyclohexane	50	0	55.8	ug/L	112				72	132	
Benzene	50	0	52.3	ug/L	105				79	120	
1,2-Dichloroethane	50	0	58.4	ug/L	117				73	128	
Trichloroethene	50	140	180	ug/L	80				79	123	
1,2-Dichloropropane	50	0	50.4	ug/L	101				78	122	
Bromodichloromethane	50	0	54.7	ug/L	109				79	125	
4-Methyl-2-Pentanone	250	0	280	ug/L	112				67	130	
Toluene	50	0	53.4	ug/L	107				80	121	
t-1,3-Dichloropropene	50	0	54.4	ug/L	109				73	127	
cis-1,3-Dichloropropene	50	0	53.1	ug/L	106				75	124	
1,1,2-Trichloroethane	50	0	52.9	ug/L	106				80	119	
2-Hexanone	250	0	300	ug/L	120				57	139	
Dibromochloromethane	50	0	54.1	ug/L	108				74	126	
1,2-Dibromoethane	50	0	53.8	ug/L	108				77	121	
Tetrachloroethene	50	4.50	58.2	ug/L	107				74	129	
Chlorobenzene	50	0	51.8	ug/L	104				82	118	
Ethyl Benzene	50	0	54.0	ug/L	108				79	121	
m/p-Xylenes	100	0	110	ug/L	110				80	121	
o-Xylene	50	0	52.0	ug/L	104				78	122	
Styrene	50	0	55.7	ug/L	111				78	123	
Bromoform	50	0	54.7	ug/L	109				66	130	
Isopropylbenzene	50	0	51.7	ug/L	103				72	131	
1,1,2,2-Tetrachloroethane	50	0	50.4	ug/L	101				71	121	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1608

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Parameter	Spike	Sample	Result	Units	Rec			RPD	Limits		
		Result			Rec	Qual	RPD	Qual	Low	High	RPD
1,3-Dichlorobenzene	50	0	51.3	ug/L	103				80	119	
1,4-Dichlorobenzene	50	0	50.0	ug/L	100				79	118	
1,2-Dichlorobenzene	50	0	51.6	ug/L	103				80	119	
1,2-Dibromo-3-Chloropropane	50	0	56.1	ug/L	112				62	128	
1,2,4-Trichlorobenzene	50	0	50.9	ug/L	102				69	130	
1,2,3-Trichlorobenzene	50	0	49.8	ug/L	100				69	129	