

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q1608 SAS No.: Q1608 SDG NO.: Q1608
 Lab File ID: VX045431.D Date Analyzed: 03/25/2025
 Instrument ID: MSVOA_X Time Analyzed: 11:31
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	83260	5.54	144945	6.76	129576	10.05
UPPER LIMIT	166520	6.044	289890	7.257	259152	10.549
LOWER LIMIT	41630	5.044	72472.5	6.257	64788	9.549
EPA SAMPLE NO.						
DUP08-20250318DL	66948	5.54	133107	6.76	120920	10.05
RE132D1-20250318	72182	5.54	140177	6.76	126304	10.05
RE132D1-20250318MS	75274	5.54	137526	6.76	121362	10.05
RE132D1-20250318MSD	80493	5.54	144565	6.76	128766	10.05
RE132D2-20250318RE	81844	5.54	157686	6.76	147292	10.05
RE132D4-20250318DL	69510	5.54	138470	6.76	127707	10.05
RE132D5-20250318DL	67394	5.54	132021	6.76	120711	10.05
RE132D6-20250318DL	64442	5.54	126973	6.76	112628	10.05
RE132D7-20250318DL	72406	5.54	142524	6.76	130699	10.05
TT101D2-20250319DL	70652	5.54	139172	6.76	131487	10.05
VX0325WBL01	71237	5.55	137150	6.76	123936	10.05
VX0325WBS02	88260	5.54	159282	6.76	137098	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.