

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q1608 SAS No.: Q1608 SDG NO.: Q1608
 Lab File ID: VX045375.D Date Analyzed: 03/21/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:00
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	63153	12.018				
UPPER LIMIT	126306	12.518				
LOWER LIMIT	31576.5	11.518				
EPA SAMPLE NO.						
TT172S1-20250317	51514	12.02				
RE139D1-20250317	48854	12.02				
RE139D2-20250317	50801	12.02				
TT205S1-20250317	44710	12.02				
DUP07-20250317	50856	12.02				
TT163S1-20250317	44709	12.02				
RW11-MW01S-20250317	47301	12.02				
VX0321WBL01	52894	12.02				
VX0321WBS01	67294	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.