

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q1608 SAS No.: Q1608 SDG NO.: Q1608
 Lab File ID: VX045403.D Date Analyzed: 03/24/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:26
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	64628	12.018				
UPPER LIMIT	129256	12.518				
LOWER LIMIT	32314	11.518				
EPA SAMPLE NO.						
RE139D1-20250317DL	49744	12.02				
RE139D2-20250317DL	46305	12.02				
RW11-MW01I-20250317	47787	12.02				
GM38-RW3-MW1-20250318	52056	12.02				
DUP08-20250318	45448	12.02				
RE132D2-20250318	48538	12.02				
RE132D3-20250318	47134	12.02				
RE132D4-20250318	54723	12.02				
RE132D5-20250318	50372	12.02				
RE132D6-20250318	50651	12.02				
RE132D7-20250318	49435	12.02				
RE117D2-20250319	50122	12.02				
TT101D2-20250319	47597	12.02				
TT161S1-20250319	49396	12.02				
TT189D2-20250319	45637	12.02				
RW8-MW01S-20250319	50743	12.02				
RW8-MW01D1-20250319	49536	12.02				
RW8-MW01D2-20250319	56752	12.02				
RW8-MW01D3-20250319	46255	12.02				
DUP09-20250319	42476	12.02				
TB	52355	12.02				
VX0324WBL01	49419	12.02				
VX0324WBS01	63508	12.02				

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q1608 SAS No.: Q1608 SDG NO.: Q1608
 Lab File ID: VX045403.D Date Analyzed: 03/24/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:26
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	64628	12.018				
UPPER LIMIT	129256	12.518				
LOWER LIMIT	32314	11.518				
EPA SAMPLE NO.						

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.