

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135116

Slope : 99.2

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/20/2025	16:14
2	CAL2	1	Water	NA	NA	20.2	7.01	03/20/2025	16:15
3	CAL3	1	Water	NA	NA	20.3	10.02	03/20/2025	16:20
4	ICV	1	Water	NA	NA	20.2	7.00	03/20/2025	16:22
5	CCV1	1	Water	NA	NA	20.1	2.01	03/20/2025	16:25
6	Q1609-02	1	Solid	20.02	20	23.4	8.85	03/20/2025	17:00
7	Q1609-02DUP	1	Solid	20.03	20	23.5	8.86	03/20/2025	17:01
8	CCV2	1	Water	NA	NA	20.3	12.02	03/21/2025	17:11

WORKLIST(Hardcopy Internal Chain)

VB 135116

WorkList Name : ph s q1609

WorkList ID : 188410

Department : Wet-Chemistry

Date : 03-20-2025 13:09:55

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1609-02	WC-SCRN-01-C	Solid	pH	Cool 4 deg C	ENTA05	I51	03/19/2025	9045D

Date/Time 03/20/25 16:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 03/20/25 18:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]