

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: Q1609

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135117

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	03/20/2025
Sample ID: CCB1 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	03/20/2025
Sample ID: CCB2 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	03/20/2025
Sample ID: CCB3 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	03/20/2025

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: Q1609

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135160

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB ASTM COD	mg/L	< 5.0000	5.0000	U	1.50	10	01/22/2025
Sample ID: CCB1 ASTM COD	mg/L	< 5.0000	5.0000	U	1.50	10	03/24/2025
Sample ID: CCB2 ASTM COD	mg/L	< 5.0000	5.0000	U	1.50	10	03/24/2025

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: Q1609

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135188

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 ASTM Ammonia	mg/L	< 0.0500	0.0500	U	0.030	0.1	03/26/2025
Sample ID: CCB1 ASTM Ammonia	mg/L	< 0.0500	0.0500	U	0.030	0.1	03/26/2025
Sample ID: CCB2 ASTM Ammonia	mg/L	< 0.0500	0.0500	U	0.030	0.1	03/26/2025

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: Q1609

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135188

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
---------	-------	--------	----------------------	--------------	-----	-----	------------------