

Matrix Spike Summary

Client:	ENTACT	SDG No.:	Q1609
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q1609-04
Client ID:	WC-SCRN-01-CMS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
ASTM COD	mg/L	75-125	57.2		17.7		50.0	1	79		03/24/2025
ASTM Ammonia	mg/L	75-125	0.98		0.030	U	1	1	98		03/26/2025
ASTM Oil and Grease	mg/L	78-114	20.4		0.30	J	20.0	1	101		03/24/2025

Matrix Spike Summary

Client:	ENTACT	SDG No.:	Q1609
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q1609-04
Client ID:	WC-SCRN-01-CMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
ASTM COD	mg/L	75-125	59.2		17.7		50.0	1	83		03/24/2025
ASTM Ammonia	mg/L	75-125	0.99		0.030	U	1	1	99		03/26/2025
ASTM Oil and Grease	mg/L	78-114	20.6		0.30	J	20.0	1	102		03/24/2025

Matrix Spike Summary

Client:	ENTACT	SDG No.:	Q1609
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q1632-01
Client ID:	PIER-1-2MS	Percent Solids for Spike Sample:	95.4

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/Kg	75-125	36.6		20.9	J	105	1	15	*	03/25/2025

Matrix Spike Summary

Client:	ENTACT	SDG No.:	Q1609
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q1632-01
Client ID:	PIER-1-2MSD	Percent Solids for Spike Sample:	95.4

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/Kg	75-125	52.3		20.9	J	105	1	30	*	03/25/2025