

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q1609

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB135150

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135150BS							
ASTM Oil and Grease	mg/L	20.0	16.7		84	1	78-114	03/24/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q1609

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB135160

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135160BS							
ASTM COD	mg/L	50	50.3		101	1	90-110	03/24/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q1609

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB135168

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135168BS							
Oil and Grease	mg/Kg	100	94.9		95	1	80-120	03/25/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q1609

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB135188

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167282BS							
ASTM Ammonia	mg/L	1	1.00		100	1	90-110	03/26/2025