

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT **SDG No.:** Q1609
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1609 **SAS No.:** Q1609
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV57	Mercury	3.72	4.0	93	90 - 110	CV	03/24/2025	08:25	LB135144

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT **SDG No.:** Q1609
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1609 **SAS No.:** Q1609
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV49	Mercury	5.05	5.0	101	90 - 110	CV	03/24/2025	08:30	LB135144
CCV50	Mercury	4.84	5.0	97	90 - 110	CV	03/24/2025	09:01	LB135144
CCV51	Mercury	4.77	5.0	95	90 - 110	CV	03/24/2025	09:28	LB135144
CCV52	Mercury	4.71	5.0	94	90 - 110	CV	03/24/2025	09:53	LB135144
CCV53	Mercury	4.80	5.0	96	90 - 110	CV	03/24/2025	10:09	LB135144



- 2a -

Client:	<u>ENTACT</u>	SDG No.:	<u>Q1609</u>	
Contract:	<u>ENTA05</u>	Lab Code:	<u>CHEM</u>	Case No.: <u>Q1609</u> SAS No.: <u>Q1609</u>
Initial Calibration Source:	<u>EPA</u>			
Continuing Calibration Source:	Inorganic Ventures			

Sample ID	Analyte	Result	True Value	%	Acceptance	M	Analysis Date	Analysis Time	Run Number
		ug/L		Recovery	Window (%R)				
ICV01	Arsenic	999	1000	100	90 - 110	P	03/24/2025	10:33	LB135158
	Barium	482	520	93	90 - 110	P	03/24/2025	10:33	LB135158
	Cadmium	494	510	97	90 - 110	P	03/24/2025	10:33	LB135158
	Chromium	529	520	102	90 - 110	P	03/24/2025	10:33	LB135158
	Copper	532	510	104	90 - 110	P	03/24/2025	10:33	LB135158
	Lead	982	1000	98	90 - 110	P	03/24/2025	10:33	LB135158
	Nickel	506	530	95	90 - 110	P	03/24/2025	10:33	LB135158
	Selenium	1030	1000	103	90 - 110	P	03/24/2025	10:33	LB135158
	Silver	258	250	103	90 - 110	P	03/24/2025	10:33	LB135158
	Zinc	1040	1000	104	90 - 110	P	03/24/2025	10:33	LB135158

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT SDG No.: Q1609
Contract: ENTA05 Lab Code: CHEM Case No.: Q1609 SAS No.: Q1609
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	18.2	20.0	91	80 - 120	P	03/24/2025	10:45	LB135158
	Barium	89.4	100	89	80 - 120	P	03/24/2025	10:45	LB135158
	Cadmium	5.64	6.0	94	80 - 120	P	03/24/2025	10:45	LB135158
	Chromium	9.80	10.0	98	80 - 120	P	03/24/2025	10:45	LB135158
	Copper	22.5	20.0	112	80 - 120	P	03/24/2025	10:45	LB135158
	Lead	11.7	12.0	97	80 - 120	P	03/24/2025	10:45	LB135158
	Nickel	38.5	40.0	96	80 - 120	P	03/24/2025	10:45	LB135158
	Selenium	18.3	20.0	92	80 - 120	P	03/24/2025	10:45	LB135158
	Silver	10.6	10.0	106	80 - 120	P	03/24/2025	10:45	LB135158
	Zinc	41.9	40.0	105	80 - 120	P	03/24/2025	10:45	LB135158

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT **SDG No.:** Q1609
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1609 **SAS No.:** Q1609
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5100	5000	102	90 - 110	P	03/24/2025	11:21	LB135158
	Barium	9430	10000	94	90 - 110	P	03/24/2025	11:21	LB135158
	Cadmium	2440	2500	98	90 - 110	P	03/24/2025	11:21	LB135158
	Chromium	1020	1000	102	90 - 110	P	03/24/2025	11:21	LB135158
	Copper	1270	1250	102	90 - 110	P	03/24/2025	11:21	LB135158
	Lead	4920	5000	98	90 - 110	P	03/24/2025	11:21	LB135158
	Nickel	2440	2500	98	90 - 110	P	03/24/2025	11:21	LB135158
	Selenium	5080	5000	102	90 - 110	P	03/24/2025	11:21	LB135158
	Silver	1260	1250	101	90 - 110	P	03/24/2025	11:21	LB135158
	Zinc	2540	2500	102	90 - 110	P	03/24/2025	11:21	LB135158
CCV02	Arsenic	5010	5000	100	90 - 110	P	03/24/2025	12:12	LB135158
	Barium	9560	10000	96	90 - 110	P	03/24/2025	12:12	LB135158
	Cadmium	2400	2500	96	90 - 110	P	03/24/2025	12:12	LB135158
	Chromium	982	1000	98	90 - 110	P	03/24/2025	12:12	LB135158
	Copper	1260	1250	101	90 - 110	P	03/24/2025	12:12	LB135158
	Lead	4830	5000	97	90 - 110	P	03/24/2025	12:12	LB135158
	Nickel	2420	2500	97	90 - 110	P	03/24/2025	12:12	LB135158
	Selenium	5070	5000	101	90 - 110	P	03/24/2025	12:12	LB135158
	Silver	1230	1250	98	90 - 110	P	03/24/2025	12:12	LB135158
	Zinc	2490	2500	100	90 - 110	P	03/24/2025	12:12	LB135158
CCV03	Arsenic	4970	5000	99	90 - 110	P	03/24/2025	13:07	LB135158
	Barium	9510	10000	95	90 - 110	P	03/24/2025	13:07	LB135158
	Cadmium	2380	2500	95	90 - 110	P	03/24/2025	13:07	LB135158
	Chromium	980	1000	98	90 - 110	P	03/24/2025	13:07	LB135158
	Copper	1250	1250	100	90 - 110	P	03/24/2025	13:07	LB135158
	Lead	4800	5000	96	90 - 110	P	03/24/2025	13:07	LB135158
	Nickel	2390	2500	95	90 - 110	P	03/24/2025	13:07	LB135158
	Selenium	4990	5000	100	90 - 110	P	03/24/2025	13:07	LB135158
	Silver	1220	1250	98	90 - 110	P	03/24/2025	13:07	LB135158
	Zinc	2470	2500	99	90 - 110	P	03/24/2025	13:07	LB135158
CCV04	Arsenic	4830	5000	96	90 - 110	P	03/24/2025	14:06	LB135158
	Barium	9200	10000	92	90 - 110	P	03/24/2025	14:06	LB135158
	Cadmium	2330	2500	93	90 - 110	P	03/24/2025	14:06	LB135158
	Chromium	951	1000	95	90 - 110	P	03/24/2025	14:06	LB135158

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT **SDG No.:** Q1609

Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1609 **SAS No.:** Q1609

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Copper	1230	1250	98	90 - 110	P	03/24/2025	14:06	LB135158
	Lead	4680	5000	94	90 - 110	P	03/24/2025	14:06	LB135158
	Nickel	2360	2500	94	90 - 110	P	03/24/2025	14:06	LB135158
	Selenium	4930	5000	99	90 - 110	P	03/24/2025	14:06	LB135158
	Silver	1190	1250	95	90 - 110	P	03/24/2025	14:06	LB135158
	Zinc	2430	2500	97	90 - 110	P	03/24/2025	14:06	LB135158
CCV05	Arsenic	4860	5000	97	90 - 110	P	03/24/2025	15:30	LB135158
	Barium	9350	10000	94	90 - 110	P	03/24/2025	15:30	LB135158
	Cadmium	2310	2500	92	90 - 110	P	03/24/2025	15:30	LB135158
	Chromium	940	1000	94	90 - 110	P	03/24/2025	15:30	LB135158
	Copper	1220	1250	98	90 - 110	P	03/24/2025	15:30	LB135158
	Lead	4640	5000	93	90 - 110	P	03/24/2025	15:30	LB135158
	Nickel	2330	2500	93	90 - 110	P	03/24/2025	15:30	LB135158
	Selenium	4950	5000	99	90 - 110	P	03/24/2025	15:30	LB135158
	Silver	1180	1250	94	90 - 110	P	03/24/2025	15:30	LB135158
	Zinc	2400	2500	96	90 - 110	P	03/24/2025	15:30	LB135158
CCV06	Arsenic	4700	5000	94	90 - 110	P	03/24/2025	16:39	LB135158
	Barium	9200	10000	92	90 - 110	P	03/24/2025	16:39	LB135158
	Cadmium	2340	2500	94	90 - 110	P	03/24/2025	16:39	LB135158
	Chromium	920	1000	92	90 - 110	P	03/24/2025	16:39	LB135158
	Copper	1190	1250	95	90 - 110	P	03/24/2025	16:39	LB135158
	Lead	4510	5000	90	90 - 110	P	03/24/2025	16:39	LB135158
	Nickel	2260	2500	90	90 - 110	P	03/24/2025	16:39	LB135158
	Selenium	4790	5000	96	90 - 110	P	03/24/2025	16:39	LB135158
	Silver	1170	1250	93	90 - 110	P	03/24/2025	16:39	LB135158
	Zinc	2370	2500	95	90 - 110	P	03/24/2025	16:39	LB135158
	Arsenic	4740	5000	95	90 - 110	P	03/24/2025	17:29	LB135158
	Barium	9010	10000	90	90 - 110	P	03/24/2025	17:29	LB135158
CCV07	Cadmium	2260	2500	90	90 - 110	P	03/24/2025	17:29	LB135158
	Chromium	938	1000	94	90 - 110	P	03/24/2025	17:29	LB135158
	Copper	1190	1250	96	90 - 110	P	03/24/2025	17:29	LB135158
	Lead	4550	5000	91	90 - 110	P	03/24/2025	17:29	LB135158
	Nickel	2280	2500	91	90 - 110	P	03/24/2025	17:29	LB135158
	Selenium	4830	5000	97	90 - 110	P	03/24/2025	17:29	LB135158

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT **SDG No.:** Q1609
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1609 **SAS No.:** Q1609
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Silver	1180	1250	94	90 - 110	P	03/24/2025	17:29	LB135158
	Zinc	2400	2500	96	90 - 110	P	03/24/2025	17:29	LB135158
CCV08	Arsenic	4760	5000	95	90 - 110	P	03/24/2025	18:01	LB135158
	Barium	9090	10000	91	90 - 110	P	03/24/2025	18:01	LB135158
	Cadmium	2250	2500	90	90 - 110	P	03/24/2025	18:01	LB135158
	Chromium	931	1000	93	90 - 110	P	03/24/2025	18:01	LB135158
	Copper	1200	1250	96	90 - 110	P	03/24/2025	18:01	LB135158
	Lead	4550	5000	91	90 - 110	P	03/24/2025	18:01	LB135158
	Nickel	2270	2500	91	90 - 110	P	03/24/2025	18:01	LB135158
	Selenium	4850	5000	97	90 - 110	P	03/24/2025	18:01	LB135158
	Silver	1170	1250	94	90 - 110	P	03/24/2025	18:01	LB135158
	Zinc	2340	2500	94	90 - 110	P	03/24/2025	18:01	LB135158

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT SDG No.: Q1609
 Contract: ENTA05 Lab Code: CHEM Case No.: Q1609 SAS No.: Q1609
 Initial Calibration Source: EPA
 Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	989	1000	99	90 - 110	P	03/25/2025	10:03	LB135182
	Barium	493	520	95	90 - 110	P	03/25/2025	10:03	LB135182
	Cadmium	494	510	97	90 - 110	P	03/25/2025	10:03	LB135182
	Chromium	518	520	100	90 - 110	P	03/25/2025	10:03	LB135182
	Copper	526	510	103	90 - 110	P	03/25/2025	10:03	LB135182
	Lead	975	1000	98	90 - 110	P	03/25/2025	10:03	LB135182
	Nickel	508	530	96	90 - 110	P	03/25/2025	10:03	LB135182
	Selenium	1020	1000	102	90 - 110	P	03/25/2025	10:03	LB135182
	Silver	253	250	101	90 - 110	P	03/25/2025	10:03	LB135182
	Zinc	1010	1000	101	90 - 110	P	03/25/2025	10:03	LB135182



- 2a -

Client:	<u>ENTACT</u>	SDG No.:	<u>Q1609</u>	
Contract:	<u>ENTA05</u>	Lab Code:	<u>CHEM</u>	Case No.: <u>Q1609</u> SAS No.: <u>Q1609</u>
Initial Calibration Source:	<u>EPA</u>			
Continuing Calibration Source:	Inorganic Ventures			

Sample ID	Analyte	Result	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
		ug/L							
LLICV01	Arsenic	18.5	20.0	93	80 - 120	P	03/25/2025	10:14	LB135182
	Barium	91.1	100	91	80 - 120	P	03/25/2025	10:14	LB135182
	Cadmium	5.74	6.0	96	80 - 120	P	03/25/2025	10:14	LB135182
	Chromium	9.50	10.0	95	80 - 120	P	03/25/2025	10:14	LB135182
	Copper	22.7	20.0	114	80 - 120	P	03/25/2025	10:14	LB135182
	Lead	11.2	12.0	93	80 - 120	P	03/25/2025	10:14	LB135182
	Nickel	39.0	40.0	98	80 - 120	P	03/25/2025	10:14	LB135182
	Selenium	21.7	20.0	108	80 - 120	P	03/25/2025	10:14	LB135182
	Silver	10.9	10.0	109	80 - 120	P	03/25/2025	10:14	LB135182
	Zinc	43.0	40.0	108	80 - 120	P	03/25/2025	10:14	LB135182

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT **SDG No.:** Q1609
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1609 **SAS No.:** Q1609
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5140	5000	103	90 - 110	P	03/25/2025	10:35	LB135182
	Barium	9690	10000	97	90 - 110	P	03/25/2025	10:35	LB135182
	Cadmium	2470	2500	99	90 - 110	P	03/25/2025	10:35	LB135182
	Chromium	1010	1000	101	90 - 110	P	03/25/2025	10:35	LB135182
	Copper	1280	1250	102	90 - 110	P	03/25/2025	10:35	LB135182
	Lead	4980	5000	100	90 - 110	P	03/25/2025	10:35	LB135182
	Nickel	2460	2500	98	90 - 110	P	03/25/2025	10:35	LB135182
	Selenium	5200	5000	104	90 - 110	P	03/25/2025	10:35	LB135182
	Silver	1260	1250	101	90 - 110	P	03/25/2025	10:35	LB135182
	Zinc	2550	2500	102	90 - 110	P	03/25/2025	10:35	LB135182
CCV02	Arsenic	4950	5000	99	90 - 110	P	03/25/2025	11:27	LB135182
	Barium	9380	10000	94	90 - 110	P	03/25/2025	11:27	LB135182
	Cadmium	2390	2500	96	90 - 110	P	03/25/2025	11:27	LB135182
	Chromium	987	1000	99	90 - 110	P	03/25/2025	11:27	LB135182
	Copper	1250	1250	100	90 - 110	P	03/25/2025	11:27	LB135182
	Lead	4810	5000	96	90 - 110	P	03/25/2025	11:27	LB135182
	Nickel	2390	2500	96	90 - 110	P	03/25/2025	11:27	LB135182
	Selenium	4980	5000	100	90 - 110	P	03/25/2025	11:27	LB135182
	Silver	1230	1250	99	90 - 110	P	03/25/2025	11:27	LB135182
	Zinc	2490	2500	99	90 - 110	P	03/25/2025	11:27	LB135182
CCV03	Arsenic	5010	5000	100	90 - 110	P	03/25/2025	12:29	LB135182
	Barium	9540	10000	95	90 - 110	P	03/25/2025	12:29	LB135182
	Cadmium	2390	2500	96	90 - 110	P	03/25/2025	12:29	LB135182
	Chromium	989	1000	99	90 - 110	P	03/25/2025	12:29	LB135182
	Copper	1260	1250	101	90 - 110	P	03/25/2025	12:29	LB135182
	Lead	4830	5000	97	90 - 110	P	03/25/2025	12:29	LB135182
	Nickel	2400	2500	96	90 - 110	P	03/25/2025	12:29	LB135182
	Selenium	5070	5000	101	90 - 110	P	03/25/2025	12:29	LB135182
	Silver	1240	1250	99	90 - 110	P	03/25/2025	12:29	LB135182
	Zinc	2510	2500	100	90 - 110	P	03/25/2025	12:29	LB135182
CCV04	Arsenic	5180	5000	104	90 - 110	P	03/25/2025	13:33	LB135182
	Barium	9950	10000	100	90 - 110	P	03/25/2025	13:33	LB135182
	Cadmium	2450	2500	98	90 - 110	P	03/25/2025	13:33	LB135182
	Chromium	1000	1000	100	90 - 110	P	03/25/2025	13:33	LB135182

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT **SDG No.:** Q1609

Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1609 **SAS No.:** Q1609

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Copper	1300	1250	104	90 - 110	P	03/25/2025	13:33	LB135182
	Lead	4940	5000	99	90 - 110	P	03/25/2025	13:33	LB135182
	Nickel	2460	2500	98	90 - 110	P	03/25/2025	13:33	LB135182
	Selenium	5260	5000	105	90 - 110	P	03/25/2025	13:33	LB135182
	Silver	1270	1250	102	90 - 110	P	03/25/2025	13:33	LB135182
	Zinc	2590	2500	103	90 - 110	P	03/25/2025	13:33	LB135182
CCV05	Arsenic	5180	5000	104	90 - 110	P	03/25/2025	14:38	LB135182
	Barium	9740	10000	97	90 - 110	P	03/25/2025	14:38	LB135182
	Cadmium	2420	2500	97	90 - 110	P	03/25/2025	14:38	LB135182
	Chromium	1000	1000	100	90 - 110	P	03/25/2025	14:38	LB135182
	Copper	1280	1250	103	90 - 110	P	03/25/2025	14:38	LB135182
	Lead	4900	5000	98	90 - 110	P	03/25/2025	14:38	LB135182
	Nickel	2430	2500	97	90 - 110	P	03/25/2025	14:38	LB135182
	Selenium	5260	5000	105	90 - 110	P	03/25/2025	14:38	LB135182
	Silver	1260	1250	101	90 - 110	P	03/25/2025	14:38	LB135182
	Zinc	2560	2500	103	90 - 110	P	03/25/2025	14:38	LB135182
CCV06	Arsenic	5190	5000	104	90 - 110	P	03/25/2025	15:46	LB135182
	Barium	9910	10000	99	90 - 110	P	03/25/2025	15:46	LB135182
	Cadmium	2440	2500	98	90 - 110	P	03/25/2025	15:46	LB135182
	Chromium	1000	1000	100	90 - 110	P	03/25/2025	15:46	LB135182
	Copper	1300	1250	104	90 - 110	P	03/25/2025	15:46	LB135182
	Lead	4930	5000	98	90 - 110	P	03/25/2025	15:46	LB135182
	Nickel	2450	2500	98	90 - 110	P	03/25/2025	15:46	LB135182
	Selenium	5270	5000	105	90 - 110	P	03/25/2025	15:46	LB135182
	Silver	1260	1250	101	90 - 110	P	03/25/2025	15:46	LB135182
	Zinc	2570	2500	103	90 - 110	P	03/25/2025	15:46	LB135182
CCV07	Arsenic	4970	5000	99	90 - 110	P	03/25/2025	16:40	LB135182
	Barium	9510	10000	95	90 - 110	P	03/25/2025	16:40	LB135182
	Cadmium	2380	2500	95	90 - 110	P	03/25/2025	16:40	LB135182
	Chromium	972	1000	97	90 - 110	P	03/25/2025	16:40	LB135182
	Copper	1250	1250	100	90 - 110	P	03/25/2025	16:40	LB135182
	Lead	4770	5000	96	90 - 110	P	03/25/2025	16:40	LB135182
	Nickel	2400	2500	96	90 - 110	P	03/25/2025	16:40	LB135182
	Selenium	5020	5000	100	90 - 110	P	03/25/2025	16:40	LB135182

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT SDG No.: Q1609
Contract: ENTA05 Lab Code: CHEM Case No.: Q1609 SAS No.: Q1609
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Silver	1220	1250	98	90 - 110	P	03/25/2025	16:40	LB135182
	Zinc	2480	2500	99	90 - 110	P	03/25/2025	16:40	LB135182
CCV08	Arsenic	5000	5000	100	90 - 110	P	03/25/2025	17:01	LB135182
	Barium	9410	10000	94	90 - 110	P	03/25/2025	17:01	LB135182
	Cadmium	2410	2500	96	90 - 110	P	03/25/2025	17:01	LB135182
	Chromium	994	1000	99	90 - 110	P	03/25/2025	17:01	LB135182
	Copper	1260	1250	101	90 - 110	P	03/25/2025	17:01	LB135182
	Lead	4840	5000	97	90 - 110	P	03/25/2025	17:01	LB135182
	Nickel	2430	2500	97	90 - 110	P	03/25/2025	17:01	LB135182
	Selenium	5050	5000	101	90 - 110	P	03/25/2025	17:01	LB135182
	Silver	1240	1250	99	90 - 110	P	03/25/2025	17:01	LB135182
	Zinc	2500	2500	100	90 - 110	P	03/25/2025	17:01	LB135182