



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT **SDG No.:** Q1609
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1609 **SAS No.:** Q1609
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.19	0.2	96	40 - 160	CV	03/24/2025	08:35	LB135144
CRI01	Arsenic	18.6	20.0	93	40 - 160	P	03/24/2025	10:54	LB135158
	Barium	86.7	100	87	40 - 160	P	03/24/2025	10:54	LB135158
	Cadmium	5.57	6.0	93	40 - 160	P	03/24/2025	10:54	LB135158
	Chromium	9.94	10.0	99	40 - 160	P	03/24/2025	10:54	LB135158
	Copper	22.5	20.0	113	40 - 160	P	03/24/2025	10:54	LB135158
	Lead	12.7	12.0	106	40 - 160	P	03/24/2025	10:54	LB135158
	Nickel	38.1	40.0	95	40 - 160	P	03/24/2025	10:54	LB135158
	Selenium	19.2	20.0	96	40 - 160	P	03/24/2025	10:54	LB135158
	Silver	10.6	10.0	106	40 - 160	P	03/24/2025	10:54	LB135158
	Zinc	41.8	40.0	104	40 - 160	P	03/24/2025	10:54	LB135158
CRI01	Arsenic	21.2	20.0	106	40 - 160	P	03/25/2025	10:22	LB135182
	Barium	91.1	100	91	40 - 160	P	03/25/2025	10:22	LB135182
	Cadmium	5.71	6.0	95	40 - 160	P	03/25/2025	10:22	LB135182
	Chromium	9.81	10.0	98	40 - 160	P	03/25/2025	10:22	LB135182
	Copper	22.2	20.0	111	40 - 160	P	03/25/2025	10:22	LB135182
	Lead	11.8	12.0	98	40 - 160	P	03/25/2025	10:22	LB135182
	Nickel	39.0	40.0	98	40 - 160	P	03/25/2025	10:22	LB135182
	Selenium	21.9	20.0	110	40 - 160	P	03/25/2025	10:22	LB135182
	Silver	11.2	10.0	112	40 - 160	P	03/25/2025	10:22	LB135182
	Zinc	42.3	40.0	106	40 - 160	P	03/25/2025	10:22	LB135182