

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: ENTACT **SDG No.:** Q1609
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1609 **SAS No.:** Q1609
Matrix: Water **Level:** LOW **Client ID:** TP-5A
Sample ID: Q1619-14 **Spiked ID:** Q1619-14A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Barium	ug/L	75 - 125	5060		3670		1000	139		P
Zinc	ug/L	75 - 125	3410		2200		1000	121		P