

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q1609
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1609 **SAS No.:** Q1609

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167260BS							
Arsenic	ug/L	4000	3850		96	80 - 120	P
Barium	ug/L	1000	859		86	80 - 120	P
Cadmium	ug/L	1000	911		91	80 - 120	P
Chromium	ug/L	2000	1990		100	80 - 120	P
Copper	ug/L	1500	1530		102	80 - 120	P
Lead	ug/L	5000	4570		91	80 - 120	P
Nickel	ug/L	2500	2330		93	80 - 120	P
Selenium	ug/L	10000	9890		99	80 - 120	P
Silver	ug/L	380	377		99	80 - 120	P
Zinc	ug/L	1000	1020		102	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q1609
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1609 **SAS No.:** Q1609

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167263BS Mercury	ug/L	4.0	3.74		94	80 - 120	CV