

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1609 SAS No.: Q1609 SDG NO.: Q1609

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL01 Date Analyzed: 03/21/2025

Lab Sample No.: AR1660CCC500 Data File : PP070791.D Time Analyzed: 08:31

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.670	5.577	5.777	484.300	500.000	-3.1
Aroclor-1016-2	5.691	5.599	5.799	511.140	500.000	2.2
Aroclor-1016-3	5.753	5.661	5.861	488.920	500.000	-2.2
Aroclor-1016-4	5.851	5.758	5.958	478.140	500.000	-4.4
Aroclor-1016-5	6.144	6.051	6.251	444.460	500.000	-11.1
Aroclor-1260-1	7.261	7.171	7.371	567.510	500.000	13.5
Aroclor-1260-2	7.514	7.425	7.625	566.690	500.000	13.3
Aroclor-1260-3	7.873	7.783	7.983	579.030	500.000	15.8
Aroclor-1260-4	8.097	8.008	8.208	543.460	500.000	8.7
Aroclor-1260-5	8.420	8.328	8.528	535.350	500.000	7.1
Decachlorobiphenyl	10.240	10.152	10.352	51.740	50.000	3.5
Tetrachloro-m-xylene	4.517	4.424	4.624	50.070	50.000	0.1

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1609 SAS No.: Q1609 SDG NO.: Q1609

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL01 Date Analyzed: 03/21/2025

Lab Sample No.: AR1660CCC500 Data File : PP070791.D Time Analyzed: 08:31

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.908	4.817	5.017	488.230	500.000	-2.4
Aroclor-1016-2	4.926	4.836	5.036	486.440	500.000	-2.7
Aroclor-1016-3	5.104	5.013	5.213	489.940	500.000	-2.0
Aroclor-1016-4	5.145	5.054	5.254	500.380	500.000	0.1
Aroclor-1016-5	5.360	5.270	5.470	489.270	500.000	-2.1
Aroclor-1260-1	6.397	6.307	6.507	484.790	500.000	-3.0
Aroclor-1260-2	6.585	6.495	6.695	469.070	500.000	-6.2
Aroclor-1260-3	6.738	6.649	6.849	472.850	500.000	-5.4
Aroclor-1260-4	7.210	7.121	7.321	462.670	500.000	-7.5
Aroclor-1260-5	7.452	7.362	7.562	464.490	500.000	-7.1
Decachlorobiphenyl	8.868	8.781	8.981	46.450	50.000	-7.1
Tetrachloro-m-xylene	3.821	3.729	3.929	50.110	50.000	0.2