

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1609 SAS No.: Q1609 SDG NO.: Q1609

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/21/2025

Lab Sample No.: AR1660CCC500 Data File : PP070805.D Time Analyzed: 16:14

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.671	5.577	5.777	439.030	500.000	-12.2
Aroclor-1016-2	5.692	5.599	5.799	471.190	500.000	-5.8
Aroclor-1016-3	5.755	5.661	5.861	434.940	500.000	-13.0
Aroclor-1016-4	5.853	5.758	5.958	422.710	500.000	-15.5
Aroclor-1016-5	6.145	6.051	6.251	432.780	500.000	-13.4
Aroclor-1260-1	7.264	7.171	7.371	488.960	500.000	-2.2
Aroclor-1260-2	7.518	7.425	7.625	494.290	500.000	-1.1
Aroclor-1260-3	7.876	7.783	7.983	503.260	500.000	0.7
Aroclor-1260-4	8.101	8.008	8.208	464.080	500.000	-7.2
Aroclor-1260-5	8.421	8.328	8.528	455.290	500.000	-8.9
Decachlorobiphenyl	10.242	10.152	10.352	44.850	50.000	-10.3
Tetrachloro-m-xylene	4.519	4.424	4.624	50.440	50.000	0.9

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1609 SAS No.: Q1609 SDG NO.: Q1609

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/21/2025

Lab Sample No.: AR1660CCC500 Data File : PP070805.D Time Analyzed: 16:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.909	4.817	5.017	533.910	500.000	6.8
Aroclor-1016-2	4.928	4.836	5.036	516.340	500.000	3.3
Aroclor-1016-3	5.105	5.013	5.213	536.300	500.000	7.3
Aroclor-1016-4	5.147	5.054	5.254	544.580	500.000	8.9
Aroclor-1016-5	5.362	5.270	5.470	518.370	500.000	3.7
Aroclor-1260-1	6.398	6.307	6.507	453.500	500.000	-9.3
Aroclor-1260-2	6.586	6.495	6.695	437.710	500.000	-12.5
Aroclor-1260-3	6.740	6.649	6.849	425.300	500.000	-14.9
Aroclor-1260-4	7.212	7.121	7.321	400.010	500.000	-20.0
Aroclor-1260-5	7.453	7.362	7.562	394.230	500.000	-21.2
Decachlorobiphenyl	8.869	8.781	8.981	39.720	50.000	-20.6
Tetrachloro-m-xylene	3.823	3.729	3.929	51.430	50.000	2.9