



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1609

Client: ENTACT

Analytical Method: **8082A** **Datafile :** PP070798.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167252BS	AR1016	166.6	163	ug/kg	98				40 (71)	140 (120)	
	AR1260	166.6	172	ug/kg	103				40 (65)	140 (130)	