

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM **Case No.:** Q1609 **SAS No.:** Q1609 **SDG NO.:** Q1609

Continuing Calib Date: 03/27/2025 **Initial Calibration Date(s):** 03/27/2025 03/27/2025

Continuing Calib Time: 22:51 **Initial Calibration Time(s):** 10:13 11:49

GC Column: RTX-CLP **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.11	7.12	7.02	7.22	0.01
2,4-D	8.22	8.22	8.12	8.32	0.00
2,4,5-TP(Silvex)	9.08	9.08	8.98	9.18	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM **Case No.:** Q1609 **SAS No.:** Q1609 **SDG NO.:** Q1609

Continuing Calib Date: 03/27/2025 **Initial Calibration Date(s):** 03/27/2025 03/27/2025

Continuing Calib Time: 22:51 **Initial Calibration Time(s):** 10:13 11:49

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.53	7.54	7.44	7.64	0.01
2,4-D	8.75	8.75	8.65	8.85	0.01
2,4,5-TP(Silvex)	9.63	9.63	9.53	9.73	0.00