

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1609 SAS No.: Q1609 SDG NO.: Q1609

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL01 Date Analyzed: 03/27/2025

Lab Sample No.: HSTDCCC750 Data File : PS029565.D Time Analyzed: 13:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.082	8.982	9.182	707.420	712.500	-0.7
2,4-D	8.222	8.122	8.322	690.140	705.000	-2.1
2,4-DCAA	7.115	7.015	7.215	736.480	750.000	-1.8

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1609 SAS No.: Q1609 SDG NO.: Q1609

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL01 Date Analyzed: 03/27/2025

Lab Sample No.: HSTDCCC750 Data File : PS029565.D Time Analyzed: 13:13

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.633	9.534	9.734	722.720	712.500	1.4
2,4-D	8.747	8.648	8.848	702.680	705.000	-0.3
2,4-DCAA	7.535	7.437	7.637	739.820	750.000	-1.4