

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1609 SAS No.: Q1609 SDG NO.: Q1609

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL04 Date Analyzed: 03/28/2025

Lab Sample No.: HSTDCCC750 Data File : PS029600.D Time Analyzed: 04:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.082	8.982	9.182	707.180	712.500	-0.7
2,4-D	8.221	8.122	8.322	728.240	705.000	3.3
2,4-DCAA	7.116	7.015	7.215	733.820	750.000	-2.2

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1609 SAS No.: Q1609 SDG NO.: Q1609

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL04 Date Analyzed: 03/28/2025

Lab Sample No.: HSTDCCC750 Data File : PS029600.D Time Analyzed: 04:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.630	9.534	9.734	710.210	712.500	-0.3
2,4-D	8.745	8.648	8.848	692.900	705.000	-1.7
2,4-DCAA	7.535	7.437	7.637	732.700	750.000	-2.3