

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1609 SAS No.: Q1609 SDG NO.: Q1609

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/25/2025

Lab Sample No.: PSTDCCC050 Data File : PL094863.D Time Analyzed: 22:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.051	8.956	9.156	49.060	50.000	-1.9
Endrin	6.571	6.474	6.674	42.730	50.000	-14.5
gamma-BHC (Lindane)	4.325	4.227	4.427	46.840	50.000	-6.3
Heptachlor	4.912	4.815	5.015	44.480	50.000	-11.0
Heptachlor epoxide	5.680	5.583	5.783	45.420	50.000	-9.2
Methoxychlor	7.498	7.400	7.600	47.920	50.000	-4.2
Tetrachloro-m-xylene	3.536	3.438	3.638	47.390	50.000	-5.2

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1609 SAS No.: Q1609 SDG NO.: Q1609

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/25/2025

Lab Sample No.: PSTDCCC050 Data File : PL094863.D Time Analyzed: 22:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.903	7.807	8.007	49.180	50.000	-1.6
Endrin	5.630	5.534	5.734	49.920	50.000	-0.2
gamma-BHC (Lindane)	3.601	3.504	3.704	51.790	50.000	3.6
Heptachlor	3.939	3.842	4.042	49.530	50.000	-0.9
Heptachlor epoxide	4.721	4.625	4.825	50.350	50.000	0.7
Methoxychlor	6.603	6.507	6.707	49.960	50.000	-0.1
Tetrachloro-m-xylene	2.769	2.672	2.872	52.200	50.000	4.4