

Surrogate Summary

SDG No.: Q1609

Client: ENTACT

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL094566.D	PIBLK-PL094566.D	Decachlorobiphenyl	1	20	22.7	114		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	20.6	103		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	21.0	105		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	20.2	101		30 (77)	150 (126)
I.BLK-PL094849.D	PIBLK-PL094849.D	Decachlorobiphenyl	1	20	19.9	99		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	18.4	92		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	14.2	71		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	19.1	95		30 (77)	150 (126)
PB167311BL	PB167311BL	Decachlorobiphenyl	1	20	20.4	102		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	18.4	92		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	16.6	83		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	19.8	99		30 (77)	150 (126)
PB167233TB	PB167233TB	Decachlorobiphenyl	1	20	20.8	104		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	18.9	94		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	17.2	86		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	19.6	98		30 (77)	150 (126)
Q1609-03	WC-SCRN-01-C	Decachlorobiphenyl	1	20	23.3	116		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	19.7	98		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	21.6	108		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	21.1	105		30 (77)	150 (126)
Q1609-03MS	WC-SCRN-01-CMS	Decachlorobiphenyl	1	20	21.6	108		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	18.9	94		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	18.2	91		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	19.8	99		30 (77)	150 (126)
Q1609-03MSD	WC-SCRN-01-CMSD	Decachlorobiphenyl	1	20	21.6	108		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	19.0	95		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	18.3	91		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	19.9	100		30 (77)	150 (126)
I.BLK-PL094862.D	PIBLK-PL094862.D	Decachlorobiphenyl	1	20	19.9	99		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	18.1	91		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	18.4	92		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	18.6	93		30 (77)	150 (126)
I.BLK-PL094893.D	PIBLK-PL094893.D	Decachlorobiphenyl	1	20	21.5	107		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	19.2	96		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	19.9	100		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	19.7	99		30 (77)	150 (126)
PB167311BS	PB167311BS	Decachlorobiphenyl	1	20	24.6	123		30 (43)	150 (140)
		Tetrachloro-m-xylene	1	20	20.1	101		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	21.0	105		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	21.4	107		30 (77)	150 (126)
I.BLK-PL094901.D	PIBLK-PL094901.D	Decachlorobiphenyl	1	20	21.2	106		30 (43)	150 (140)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: Q1609

Client: ENTACT

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL094901.D	PIBLK-PL094901.D	Tetrachloro-m-xylene	1	20	17.9	90		30 (77)	150 (126)
		Decachlorobiphenyl	2	20	20.2	101		30 (43)	150 (140)
		Tetrachloro-m-xylene	2	20	18.9	95		30 (77)	150 (126)