

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF031025

Review By	anahy	Review On	3/11/2025 9:11:47 AM
Supervise By	Jagrut	Supervise On	3/11/2025 10:18:40 AM
SubDirectory	BF031025	HP Acquire Method	BNA_F
		HP Processing Method	bf031025

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12653,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF141896.D	10 Mar 2025 10:31		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF141897.D	10 Mar 2025 11:01		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF141898.D	10 Mar 2025 11:30	Compound#9,32,54,65 removed from 5 ppm	RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF141899.D	10 Mar 2025 12:00		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF141900.D	10 Mar 2025 12:29	Compound #54 Kept on LR, Method is good for DOD . Method failed for compound #77.	RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF141901.D	10 Mar 2025 12:58		RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF141902.D	10 Mar 2025 13:28	Not used	RC/JU	Not Ok
8	SSTDICC060	SSTDICC060	BF141903.D	10 Mar 2025 13:57	Compound#69 Removed from 60 ppm	RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF141904.D	10 Mar 2025 14:27	Compound#9,69 Removed from 80 ppm	RC/JU	Ok,M
10	SSTDICC050	SSTDICC050	BF141905.D	10 Mar 2025 15:20		RC/JU	Ok
11	SSTDICV040	ICVBF031025	BF141906.D	10 Mar 2025 15:53		RC/JU	Ok
12	PB167012BL	PB167012BL	BF141907.D	10 Mar 2025 17:09		RC/JU	Ok
13	SP6752	SP6752	BF141908.D	10 Mar 2025 17:39	8270 SPIKE-SP6752	RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF032425

Review By	anahy	Review On	3/25/2025 9:38:18 AM
Supervise By	Jagrut	Supervise On	3/25/2025 9:39:38 AM
SubDirectory	BF032425	HP Acquire Method	BNA_F
		HP Processing Method	bf031025

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12655,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142044.D	24 Mar 2025 09:43		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF142045.D	24 Mar 2025 10:17		RC/JU	Ok
3	PB167228BL	PB167228BL	BF142046.D	24 Mar 2025 10:47		RC/JU	Ok
4	Q1597-02	1-CONCRETE-SLAB	BF142047.D	24 Mar 2025 11:27		RC/JU	Ok
5	Q1597-04	2-CONCRETE-SLAB	BF142048.D	24 Mar 2025 11:56		RC/JU	Ok
6	Q1592-02MS	OILY-DEBRIS-COMP	BF142049.D	24 Mar 2025 12:27		RC/JU	Ok,M
7	Q1592-02MSD	OILY-DEBRIS-COMP	BF142050.D	24 Mar 2025 12:56		RC/JU	Ok,M
8	Q1619-13DL	TP-5DL	BF142051.D	24 Mar 2025 13:26	Need 2X Further Dilution	RC/JU	Dilution
9	Q1618-01	TP20250320-01	BF142052.D	24 Mar 2025 13:55		RC/JU	Ok
10	Q1618-02	TP20250320-02	BF142053.D	24 Mar 2025 14:25		RC/JU	Ok
11	Q1619-13DL2	TP-5DL2	BF142054.D	24 Mar 2025 14:55		RC/JU	Ok,M
12	Q1606-11	N48965	BF142055.D	24 Mar 2025 15:25	Internal Standard Fail	RC/JU	ReRun
13	Q1606-02	CHRT24743	BF142056.D	24 Mar 2025 15:55		RC/JU	Ok
14	Q1606-04	RBR251346	BF142057.D	24 Mar 2025 16:24		RC/JU	Ok
15	Q1606-06	RT4534	BF142058.D	24 Mar 2025 16:54		RC/JU	Ok
16	Q1606-08	RT3025	BF142059.D	24 Mar 2025 17:24		RC/JU	Ok
17	Q1606-10	CHRT28607	BF142060.D	24 Mar 2025 17:54		RC/JU	Ok
18	Q1610-03	SOIL-PILE	BF142061.D	24 Mar 2025 18:23		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF032425

Review By	anahy	Review On	3/25/2025 9:38:18 AM		
Supervise By	Jagrut	Supervise On	3/25/2025 9:39:38 AM		
SubDirectory	BF032425	HP Acquire Method	BNA_F	HP Processing Method	bf031025
STD. NAME		STD REF.#			
Tune/Reschk		SP6717			
Initial Calibration Stds		SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729			
CCC		SP6725			
Internal Standard/PEM		S12655,10ul/1000ul sample			
ICV/I.BLK		SP6686			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q1619-12	TP-4	BF142062.D	24 Mar 2025 18:53		RC/JU	Ok
20	Q1624-01	OK-01-03212025	BF142063.D	24 Mar 2025 19:23		RC/JU	Ok,M
21	Q1626-01	CO-32-1	BF142064.D	24 Mar 2025 19:53		RC/JU	Ok,M
22	Q1626-01MS	CO-32-1MS	BF142065.D	24 Mar 2025 20:22		RC/JU	Ok
23	Q1626-01MSD	CO-32-1MSD	BF142066.D	24 Mar 2025 20:51	Internal standard fail.	RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF032525

Review By	anahy	Review On	3/26/2025 9:09:40 AM
Supervise By	Jagrut	Supervise On	3/26/2025 4:50:57 PM
SubDirectory	BF032525	HP Acquire Method	BNA_F
		HP Processing Method	bf031025

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12656,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142067.D	25 Mar 2025 09:39		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF142068.D	25 Mar 2025 10:09		RC/JU	Ok
3	PB167274BL	PB167274BL	BF142069.D	25 Mar 2025 10:39		RC/JU	Ok
4	PB167274BS	PB167274BS	BF142070.D	25 Mar 2025 11:09		RC/JU	Ok,M
5	PB167261BL	PB167261BL	BF142071.D	25 Mar 2025 11:38		RC/JU	Ok
6	PB167261BS	PB167261BS	BF142072.D	25 Mar 2025 12:08		RC/JU	Ok,M
7	PB167193TB	PB167193TB	BF142073.D	25 Mar 2025 12:38		RC/JU	Ok
8	PB167230BL	PB167230BL	BF142074.D	25 Mar 2025 13:07		RC/JU	Ok
9	PB167230BS	PB167230BS	BF142075.D	25 Mar 2025 13:37		RC/JU	Ok,M
10	PB167230BSD	PB167230BSD	BF142076.D	25 Mar 2025 14:07		RC/JU	Ok,M
11	PB167254BL	PB167254BL	BF142077.D	25 Mar 2025 14:37		RC/JU	Ok
12	PB167254BS	PB167254BS	BF142078.D	25 Mar 2025 15:07		RC/JU	Ok,M
13	Q1619-06	TP-1	BF142079.D	25 Mar 2025 15:41		RC/JU	Ok
14	Q1619-08	TP-2	BF142080.D	25 Mar 2025 16:11		RC/JU	Ok
15	Q1619-10	TP-3	BF142081.D	25 Mar 2025 16:41		RC/JU	Ok
16	Q1619-14	TP-5	BF142082.D	25 Mar 2025 17:10		RC/JU	Ok
17	Q1609-03	WC-SCRN-01-C	BF142083.D	25 Mar 2025 17:39		RC/JU	Ok
18	Q1626-03	CO-32-1	BF142084.D	25 Mar 2025 18:09		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF032525

Review By	anahy	Review On	3/26/2025 9:09:40 AM		
Supervise By	Jagrut	Supervise On	3/26/2025 4:50:57 PM		
SubDirectory	BF032525	HP Acquire Method	BNA_F	HP Processing Method	bf031025
STD. NAME	STD REF.#				
Tune/Reschk	SP6717				
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729				
CCC	SP6725				
Internal Standard/PEM	S12656,10ul/1000ul sample				
ICV/I.BLK	SP6686				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q1627-01	GRID-LINE-1.2-NORTH	BF142085.D	25 Mar 2025 18:38		RC/JU	Ok
20	Q1627-01MS	GRID-LINE-1.2-NORTH	BF142086.D	25 Mar 2025 19:08		RC/JU	Ok,M
21	Q1627-01MSD	GRID-LINE-1.2-NORTH	BF142087.D	25 Mar 2025 19:37		RC/JU	Ok,M
22	Q1630-02	VNJ-206	BF142088.D	25 Mar 2025 20:07		RC/JU	Ok
23	Q1632-02	PIER-1-2	BF142089.D	25 Mar 2025 20:36		RC/JU	Ok,M
24	Q1636-04	WC-1	BF142090.D	25 Mar 2025 21:06		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF033125

Review By		Review On			
Supervise By		Supervise On			
SubDirectory	BF033125	HP Acquire Method	BNA_F	HP Processing Method	bf031025
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729				
CCC	SP6725				
Internal Standard/PEM	S12656,10ul/1000ul sample				
ICV/I.BLK	SP6686				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142174.D	31 Mar 2025 11:53		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF142175.D	31 Mar 2025 12:22	CCC fail high side for com. #9	RC/JU	Ok,NR
3	PB167369BL	PB167369BL	BF142176.D	31 Mar 2025 12:52		RC/JU	Ok
4	PB167369BS	PB167369BS	BF142177.D	31 Mar 2025 13:22		RC/JU	Ok,NR
5	PB167233TB	PB167233TB	BF142178.D	31 Mar 2025 13:51		RC/JU	Ok,NR
6	Q1657-01DL	72-11998DL	BF142179.D	31 Mar 2025 14:29		RC/JU	Ok,NR
7	Q1650-01DL	STOCK-PILE-BIN2DL	BF142180.D	31 Mar 2025 14:59	Internal Standard Fail	RC/JU	Not Ok
8	Q1654-01DL	RT3407DL	BF142181.D	31 Mar 2025 15:28	Internal Standard Fail, Need 10X further Dilution	RC/JU	Not Ok
9	Q1664-09	P001-BBDGA-002-01	BF142182.D	31 Mar 2025 15:58	Internal Standard Fail	RC/JU	ReRun
10	Q1654-01DL2	RT3407DL2	BF142183.D	31 Mar 2025 16:28	Internal Standard Fail	RC/JU	Not Ok
11	Q1664-11	P001-BBDGA-003-01	BF142184.D	31 Mar 2025 16:58		RC/JU	Ok,NR
12	Q1664-13	P001-BBDGA-004-01	BF142185.D	31 Mar 2025 17:27		RC/JU	Ok,NR
13	Q1664-17	P001-BBDGA-006-01	BF142186.D	31 Mar 2025 17:57	Internal Standard Fail	RC/JU	ReRun
14	Q1664-21	P001-BBDGA-008-01	BF142187.D	31 Mar 2025 18:27		RC/JU	Ok,NR
15	Q1664-01	P001-BBDGA-001-01	BF142188.D	31 Mar 2025 18:57		RC/JU	Ok,NR
16	Q1664-15	P001-BBDGA-005-01	BF142189.D	31 Mar 2025 19:26	Internal Standard Fail	RC/JU	ReRun
17	Q1664-19	P001-BBDGA-007-01	BF142190.D	31 Mar 2025 19:56		RC/JU	Ok,NR

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF033125

Review By		Review On			
Supervise By		Supervise On			
SubDirectory	BF033125	HP Acquire Method	BNA_F	HP Processing Method	bf031025
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729				
CCC	SP6725				
Internal Standard/PEM	S12656,10ul/1000ul sample				
ICV/I.BLK	SP6686				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q1664-07	P001-BBDGA-001-02	BF142191.D	31 Mar 2025 20:26	Internal Standard Fail	RC/JU	ReRun
19	Q1664-02MS	P001-BBDGA-001-01M	BF142192.D	31 Mar 2025 20:55	Internal Standard Fail	RC/JU	Not Ok
20	Q1664-03MSD	P001-BBDGA-001-01M	BF142193.D	31 Mar 2025 21:26	Internal Standard Fail	RC/JU	Not Ok
21	Q1671-01	WC-1	BF142194.D	31 Mar 2025 21:55	Internal Standard Fail	RC/JU	ReRun
22	Q1674-01	RT5358	BF142195.D	31 Mar 2025 22:25	Internal Standard Fail	RC/JU	ReRun
23	Q1671-04	WC-1	BF142196.D	31 Mar 2025 22:55	Internal Standard Fail	RC/JU	ReRun
24	Q1674-03	72-11991	BF142197.D	31 Mar 2025 23:25	Internal Standard Fail	RC/JU	ReRun

M : Manual Integration