

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1609 SAS No.: Q1609 SDG NO.: Q1609

EPA Sample No.: SSTDCCC040 Date Analyzed: 03/24/2025

Lab File ID: BF142045.D Time Analyzed: 10:17

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	161989	6.875	619308	8.16	352913	9.91
UPPER LIMIT	323978	7.375	1238620	8.657	705826	10.41
LOWER LIMIT	80994.5	6.375	309654	7.657	176457	9.41
EPA SAMPLE NO.						
01 OILY-DEBRIS-COMPMS	136574	6.88	525982	8.16	287440	9.91
02 OILY-DEBRIS-COMPMSD	136452	6.88	529624	8.16	299190	9.91

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.