

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q1609 SAS No.: Q1609 SDG No.: Q1609
 Instrument ID: MSVOA_N Calibration Date(s): 03/18/2025 03/18/2025
 Heated Purge: (Y/N) N Calibration Time(s): 11:44 14:09
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF100 = VN085996.D		RRF050 = VN085997.D		RRF020 = VN085998.D			
		RRF010 = VN085999.D		RRF005 = VN086000.D		RRF001 = VN086001.D			
COMPOUND	RRF100	RRF050	RRF020	RRF010	RRF005	RRF001	RRF	% RSD	
Vinyl Chloride	0.622	0.592	0.541	0.581	0.634	0.592	0.594	5.5	
1,1-Dichloroethene	0.567	0.516	0.481	0.545	0.550	0.415	0.512	11	
2-Butanone	0.331	0.292	0.277	0.297	0.319	0.296	0.302	6.5	
Carbon Tetrachloride	0.673	0.578	0.557	0.580	0.613	0.624	0.604	6.9	
Chloroform	1.119	1.017	1.011	1.101	1.149	1.245	1.107	7.9	
Benzene	1.610	1.393	1.348	1.386	1.453	1.466	1.443	6.5	
1,2-Dichloroethane	0.538	0.475	0.462	0.491	0.528	0.521	0.503	6.1	
Trichloroethene	0.394	0.346	0.335	0.353	0.380	0.435	0.374	10	
Tetrachloroethene	0.407	0.371	0.370	0.390	0.421	0.433	0.399	6.6	
Chlorobenzene	1.229	1.085	1.070	1.116	1.156	1.208	1.144	5.7	
1,2-Dichloroethane-d4	0.632	0.665	0.669	0.721	0.737		0.685	6.3	
Dibromofluoromethane	0.333	0.334	0.347	0.362	0.368		0.349	4.5	
Toluene-d8	1.309	1.266	1.253	1.289	1.217		1.267	2.8	
4-Bromofluorobenzene	0.514	0.466	0.447	0.440	0.391		0.452	9.8	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.