

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ENTA05

Lab Code: CHEM Case No.: Q1609 SAS No.: Q1609 SDG No.: Q1609

Instrument ID: MSVOA_N Calibration Date/Time: 03/21/2025 13:17

Lab File ID: VN086036.D Init. Calib. Date(s): 03/18/2025 03/18/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:44 14:09

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.594	0.588		-1.01	20
1,1-Dichloroethene	0.512	0.527		2.93	20
2-Butanone	0.302	0.313		3.64	20
Carbon Tetrachloride	0.604	0.657		8.77	20
Chloroform	1.107	1.162		4.97	20
Benzene	1.443	1.587		9.98	20
1,2-Dichloroethane	0.503	0.542		7.75	20
Trichloroethene	0.374	0.366		-2.14	20
Tetrachloroethene	0.399	0.397		-0.5	20
Chlorobenzene	1.144	1.174	0.3	2.62	20
1,2-Dichloroethane-d4	0.685	0.703		2.63	20
Dibromofluoromethane	0.349	0.369		5.73	20
Toluene-d8	1.267	1.419		12	20
4-Bromofluorobenzene	0.452	0.521		15.27	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.