

Surrogate Summary

SDG No.: Q1609

Client: ENTACT

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1609-01	WC-SCRN-01-G	1,2-Dichloroethane-d4	50	57.5	115	70 (74)	130 (125)
		Dibromofluoromethane	50	50.9	102	70 (75)	130 (124)
		Toluene-d8	50	50.8	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	43.0	86	70 (77)	130 (121)
VN0321WBL01	VN0321WBL01	1,2-Dichloroethane-d4	50	57.7	115	70 (74)	130 (125)
		Dibromofluoromethane	50	56.2	112	70 (75)	130 (124)
		Toluene-d8	50	48.6	97	70 (86)	130 (113)
		4-Bromofluorobenzene	50	44.4	89	70 (77)	130 (121)
VN0321WBS01	VN0321WBS01	1,2-Dichloroethane-d4	50	49.1	98	70 (74)	130 (125)
		Dibromofluoromethane	50	50.1	100	70 (75)	130 (124)
		Toluene-d8	50	50.9	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.4	101	70 (77)	130 (121)
VN0321WBSD0	VN0321WBSD01	1,2-Dichloroethane-d4	50	51.4	103	70 (74)	130 (125)
		Dibromofluoromethane	50	51.4	103	70 (75)	130 (124)
		Toluene-d8	50	51.0	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.6	103	70 (77)	130 (121)

