

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1609
Client: ENTACT
Analytical Method: SW8260D **Datafile :** VN086039.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0321WBSD01	Vinyl chloride	20	21.4	ug/L	107	7		70 (65)	130 (117)	20 (20)
	1,1-Dichloroethene	20	20.7	ug/L	104	2		70 (74)	130 (110)	20 (20)
	2-Butanone	100	110	ug/L	110	10		40 (65)	160 (122)	20 (20)
	Carbon Tetrachloride	20	21.5	ug/L	108	0		70 (77)	130 (113)	20 (20)
	Chloroform	20	21.4	ug/L	107	1		70 (79)	130 (113)	20 (20)
	Benzene	20	22.0	ug/L	110	1		70 (82)	130 (109)	20 (20)
	1,2-Dichloroethane	20	22.4	ug/L	112	6		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	20.1	ug/L	101	1		70 (77)	130 (113)	20 (20)
	Tetrachloroethene	20	20.7	ug/L	104	0		70 (67)	130 (123)	20 (20)
	Chlorobenzene	20	20.8	ug/L	104	2		70 (82)	130 (109)	20 (20)