

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q1609 SAS No.: Q1609 SDG NO.: Q1609
 Lab File ID: VN086036.D Date Analyzed: 03/21/2025
 Instrument ID: MSVOA_N Time Analyzed: 13:17
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	163728	8.22	253523	9.10	241486	11.87
UPPER LIMIT	327456	8.718	507046	9.6	482972	12.365
LOWER LIMIT	81864	7.718	126762	8.6	120743	11.365
EPA SAMPLE NO.						
WC-SCRN-01-G	140609	8.22	262572	9.10	254973	11.87
VN0321WBL01	129724	8.22	224711	9.10	210417	11.87
VN0321WBS01	165667	8.22	258178	9.10	240248	11.87
VN0321WBSD01	154418	8.22	246307	9.10	225945	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.