

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05  
 Lab Code: CHEM Case No.: Q1609 SAS No.: Q1609 SDG NO.: Q1609  
 Lab File ID: VN086036.D Date Analyzed: 03/21/2025  
 Instrument ID: MSVOA\_N Time Analyzed: 13:17  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 138904        | 13.788 |  |  |  |  |
| UPPER LIMIT    | 277808        | 14.288 |  |  |  |  |
| LOWER LIMIT    | 69452         | 13.288 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| WC-SCRN-01-G   | 98590         | 13.79  |  |  |  |  |
| VN0321WBL01    | 84947         | 13.79  |  |  |  |  |
| VN0321WBS01    | 125213        | 13.79  |  |  |  |  |
| VN0321WBSD01   | 115388        | 13.79  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.