

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1616 SAS No.: Q1616 SDG NO.: Q1616

Continuing Calib Date: 03/20/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 14:19 Initial Calibration Time(s): 15:13 23:38

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	5.67	5.68	5.58	5.78	0.01
Aroclor-1242-2 (2)	5.70	5.70	5.60	5.80	0.00
Aroclor-1242-3 (3)	5.76	5.76	5.66	5.86	0.00
Aroclor-1242-4 (4)	5.85	5.86	5.76	5.96	0.01
Aroclor-1242-5 (5)	6.58	6.59	6.49	6.69	0.01
Tetrachloro-m-xylene	4.52	4.53	4.43	4.63	0.01
Decachlorobiphenyl	10.24	10.25	10.15	10.35	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1616 SAS No.: Q1616 SDG NO.: Q1616

Continuing Calib Date: 03/20/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 14:19 Initial Calibration Time(s): 15:13 23:38

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.91	4.92	4.82	5.02	0.01
Aroclor-1242-2	(2)	4.93	4.94	4.84	5.04	0.01
Aroclor-1242-3	(3)	5.11	5.11	5.01	5.21	0.00
Aroclor-1242-4	(4)	5.19	5.20	5.10	5.30	0.01
Aroclor-1242-5	(5)	5.72	5.72	5.62	5.82	0.00
Tetrachloro-m-xylene		3.82	3.83	3.73	3.93	0.01
Decachlorobiphenyl		8.87	8.88	8.78	8.98	0.01