

### CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1616 SAS No.: Q1616 SDG NO.: Q1616

Continuing Calib Date: 03/21/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 03:09 Initial Calibration Time(s): 15:13 23:38

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1016-1       | (1) | 5.67       | 5.68      | 5.58      | 5.78  | 0.01       |
| Aroclor-1016-2       | (2) | 5.70       | 5.70      | 5.60      | 5.80  | 0.00       |
| Aroclor-1016-3       | (3) | 5.76       | 5.76      | 5.66      | 5.86  | 0.00       |
| Aroclor-1016-4       | (4) | 5.86       | 5.86      | 5.76      | 5.96  | 0.00       |
| Aroclor-1016-5       | (5) | 6.15       | 6.15      | 6.05      | 6.25  | 0.00       |
| Aroclor-1260-1       | (1) | 7.27       | 7.27      | 7.17      | 7.37  | 0.00       |
| Aroclor-1260-2       | (2) | 7.52       | 7.53      | 7.43      | 7.63  | 0.01       |
| Aroclor-1260-3       | (3) | 7.88       | 7.88      | 7.78      | 7.98  | 0.00       |
| Aroclor-1260-4       | (4) | 8.10       | 8.11      | 8.01      | 8.21  | 0.01       |
| Aroclor-1260-5       | (5) | 8.42       | 8.43      | 8.33      | 8.53  | 0.01       |
| Tetrachloro-m-xylene |     | 4.52       | 4.52      | 4.42      | 4.62  | 0.00       |
| Decachlorobiphenyl   |     | 10.24      | 10.25     | 10.15     | 10.35 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1616 SAS No.: Q1616 SDG NO.: Q1616

Continuing Calib Date: 03/21/2025 Initial Calibration Date(s): 03/11/2025 03/11/2025

Continuing Calib Time: 03:09 Initial Calibration Time(s): 15:13 23:38

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|------------------------|------|------------|
| Aroclor-1016-1       | (1) | 4.91       | 4.92      | 4.82                   | 5.02 | 0.01       |
| Aroclor-1016-2       | (2) | 4.93       | 4.94      | 4.84                   | 5.04 | 0.01       |
| Aroclor-1016-3       | (3) | 5.10       | 5.11      | 5.01                   | 5.21 | 0.01       |
| Aroclor-1016-4       | (4) | 5.15       | 5.15      | 5.05                   | 5.25 | 0.00       |
| Aroclor-1016-5       | (5) | 5.36       | 5.37      | 5.27                   | 5.47 | 0.01       |
| Aroclor-1260-1       | (1) | 6.40       | 6.41      | 6.31                   | 6.51 | 0.01       |
| Aroclor-1260-2       | (2) | 6.59       | 6.60      | 6.50                   | 6.70 | 0.01       |
| Aroclor-1260-3       | (3) | 6.74       | 6.75      | 6.65                   | 6.85 | 0.01       |
| Aroclor-1260-4       | (4) | 7.21       | 7.22      | 7.12                   | 7.32 | 0.01       |
| Aroclor-1260-5       | (5) | 7.45       | 7.46      | 7.36                   | 7.56 | 0.01       |
| Tetrachloro-m-xylene |     | 3.82       | 3.83      | 3.73                   | 3.93 | 0.01       |
| Decachlorobiphenyl   |     | 8.87       | 8.88      | 8.78                   | 8.98 | 0.01       |