

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1616 SAS No.: Q1616 SDG NO.: Q1616

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL01 Date Analyzed: 03/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP070757.D Time Analyzed: 18:07

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.674	5.577	5.777	467.710	500.000	-6.5
Aroclor-1016-2	5.696	5.599	5.799	520.200	500.000	4.0
Aroclor-1016-3	5.758	5.661	5.861	474.870	500.000	-5.0
Aroclor-1016-4	5.855	5.758	5.958	458.710	500.000	-8.3
Aroclor-1016-5	6.148	6.051	6.251	493.880	500.000	-1.2
Aroclor-1260-1	7.266	7.171	7.371	553.020	500.000	10.6
Aroclor-1260-2	7.520	7.425	7.625	555.900	500.000	11.2
Aroclor-1260-3	7.879	7.783	7.983	577.190	500.000	15.4
Aroclor-1260-4	8.103	8.008	8.208	530.420	500.000	6.1
Aroclor-1260-5	8.424	8.328	8.528	525.190	500.000	5.0
Decachlorobiphenyl	10.243	10.152	10.352	51.240	50.000	2.5
Tetrachloro-m-xylene	4.522	4.424	4.624	50.600	50.000	1.2

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1616 SAS No.: Q1616 SDG NO.: Q1616

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL01 Date Analyzed: 03/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP070757.D Time Analyzed: 18:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.910	4.817	5.017	515.250	500.000	3.1
Aroclor-1016-2	4.928	4.836	5.036	509.850	500.000	2.0
Aroclor-1016-3	5.106	5.013	5.213	514.320	500.000	2.9
Aroclor-1016-4	5.147	5.054	5.254	509.900	500.000	2.0
Aroclor-1016-5	5.362	5.270	5.470	503.690	500.000	0.7
Aroclor-1260-1	6.398	6.307	6.507	492.950	500.000	-1.4
Aroclor-1260-2	6.587	6.495	6.695	481.580	500.000	-3.7
Aroclor-1260-3	6.741	6.649	6.849	480.490	500.000	-3.9
Aroclor-1260-4	7.213	7.121	7.321	459.530	500.000	-8.1
Aroclor-1260-5	7.453	7.362	7.562	455.160	500.000	-9.0
Decachlorobiphenyl	8.870	8.781	8.981	46.140	50.000	-7.7
Tetrachloro-m-xylene	3.823	3.729	3.929	53.220	50.000	6.4