

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1616 SAS No.: Q1616 SDG NO.: Q1616

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP070771.D Time Analyzed: 22:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.673	5.577	5.777	481.020	500.000	-3.8
Aroclor-1016-2	5.695	5.599	5.799	524.190	500.000	4.8
Aroclor-1016-3	5.758	5.661	5.861	487.900	500.000	-2.4
Aroclor-1016-4	5.855	5.758	5.958	473.620	500.000	-5.3
Aroclor-1016-5	6.148	6.051	6.251	492.310	500.000	-1.5
Aroclor-1260-1	7.265	7.171	7.371	575.340	500.000	15.1
Aroclor-1260-2	7.519	7.425	7.625	576.630	500.000	15.3
Aroclor-1260-3	7.879	7.783	7.983	580.090	500.000	16.0
Aroclor-1260-4	8.103	8.008	8.208	537.760	500.000	7.6
Aroclor-1260-5	8.423	8.328	8.528	533.390	500.000	6.7
Decachlorobiphenyl	10.244	10.152	10.352	51.700	50.000	3.4
Tetrachloro-m-xylene	4.521	4.424	4.624	50.910	50.000	1.8

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1616 SAS No.: Q1616 SDG NO.: Q1616

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP070771.D Time Analyzed: 22:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.909	4.817	5.017	524.130	500.000	4.8
Aroclor-1016-2	4.928	4.836	5.036	512.670	500.000	2.5
Aroclor-1016-3	5.105	5.013	5.213	522.220	500.000	4.4
Aroclor-1016-4	5.147	5.054	5.254	512.210	500.000	2.4
Aroclor-1016-5	5.362	5.270	5.470	566.790	500.000	13.4
Aroclor-1260-1	6.399	6.307	6.507	516.620	500.000	3.3
Aroclor-1260-2	6.587	6.495	6.695	497.790	500.000	-0.4
Aroclor-1260-3	6.741	6.649	6.849	499.790	500.000	0.0
Aroclor-1260-4	7.213	7.121	7.321	473.500	500.000	-5.3
Aroclor-1260-5	7.454	7.362	7.562	469.720	500.000	-6.1
Decachlorobiphenyl	8.870	8.781	8.981	47.290	50.000	-5.4
Tetrachloro-m-xylene	3.823	3.729	3.929	54.520	50.000	9.0