

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1617 SAS No.: Q1617 SDG NO.: Q1617

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL03 Date Analyzed: 03/21/2025

Lab Sample No.: AR1660CCC500 Data File : PP070785.D Time Analyzed: 03:09

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.673	5.577	5.777	490.740	500.000	-1.9
Aroclor-1016-2	5.696	5.599	5.799	523.110	500.000	4.6
Aroclor-1016-3	5.758	5.661	5.861	495.020	500.000	-1.0
Aroclor-1016-4	5.855	5.758	5.958	481.110	500.000	-3.8
Aroclor-1016-5	6.148	6.051	6.251	504.410	500.000	0.9
Aroclor-1260-1	7.267	7.171	7.371	583.140	500.000	16.6
Aroclor-1260-2	7.519	7.425	7.625	577.450	500.000	15.5
Aroclor-1260-3	7.878	7.783	7.983	574.450	500.000	14.9
Aroclor-1260-4	8.104	8.008	8.208	557.790	500.000	11.6
Aroclor-1260-5	8.424	8.328	8.528	546.230	500.000	9.2
Decachlorobiphenyl	10.244	10.152	10.352	53.000	50.000	6.0
Tetrachloro-m-xylene	4.521	4.424	4.624	51.110	50.000	2.2

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1617 SAS No.: Q1617 SDG NO.: Q1617

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL03 Date Analyzed: 03/21/2025

Lab Sample No.: AR1660CCC500 Data File : PP070785.D Time Analyzed: 03:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.908	4.817	5.017	518.700	500.000	3.7
Aroclor-1016-2	4.927	4.836	5.036	490.990	500.000	-1.8
Aroclor-1016-3	5.104	5.013	5.213	509.770	500.000	2.0
Aroclor-1016-4	5.146	5.054	5.254	496.480	500.000	-0.7
Aroclor-1016-5	5.361	5.270	5.470	554.960	500.000	11.0
Aroclor-1260-1	6.397	6.307	6.507	510.380	500.000	2.1
Aroclor-1260-2	6.586	6.495	6.695	497.120	500.000	-0.6
Aroclor-1260-3	6.739	6.649	6.849	495.140	500.000	-1.0
Aroclor-1260-4	7.211	7.121	7.321	479.850	500.000	-4.0
Aroclor-1260-5	7.453	7.362	7.562	483.520	500.000	-3.3
Decachlorobiphenyl	8.869	8.781	8.981	49.040	50.000	-1.9
Tetrachloro-m-xylene	3.821	3.729	3.929	52.110	50.000	4.2