

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1620 SAS No.: Q1620 SDG NO.: Q1620

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 03/28/2025

Lab File ID: BN036757.D Time Analyzed: 03:46

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1974	7.703	4607	10.49	2750	14.33
UPPER LIMIT	3948	8.203	9214	10.988	5500	14.834
LOWER LIMIT	987	7.203	2303.5	9.988	1375	13.834
EPA SAMPLE NO.						
01 PB167269BL	2126	7.70	4744	10.50	2602	14.35

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.