

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1620 SAS No.: Q1620 SDG NO.: Q1620

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 03/31/2025

Lab File ID: BN036809.D Time Analyzed: 10:39

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1888	7.695	4656	10.48	2798	14.33
UPPER LIMIT	3776	8.195	9312	10.977	5596	14.834
LOWER LIMIT	944	7.195	2328	9.977	1399	13.834
EPA SAMPLE NO.						
01 PB167269BSD	2858	7.70	7352	10.48	4021	14.33

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.