

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q1620

SDG NO.: Q1620

Lab File ID: BN036808.D

DFTPP Injection Date: 03/31/2025

Instrument ID: BNA_N

DFTPP Injection Time: 10:00

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	70.2
68	Less than 2.0% of mass 69	1 (1.8) 1
69	Mass 69 relative abundance	58.4
70	Less than 2.0% of mass 69	0.4 (0.6) 1
127	10.0 - 80.0% of mass 198	54.1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7.2
275	10.0 - 60.0% of mass 198	23.4
365	Greater than 1% of mass 198	4
441	Present, but less than mass 443	8.3
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10 (21.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN036809.D	03/31/2025	10:39
PB167269BSD	PB167269BSD	BN036814.D	03/31/2025	14:16
SSTDCCC0.4EC	SSTDCCC0.4	BN036815.D	03/31/2025	15:08