

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1621 SAS No.: Q1621 SDG No.: Q1621

Instrument ID: BNA_N Calibration Date/Time: 03/29/2025 12:21

Lab File ID: BN036807.D Init. Calib. Date(s): 03/10/2025 03/10/2025

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 11:42 15:19

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.595	0.526		-11.6	50.0
Fluoranthene-d10	1.025	0.985		-3.9	50.0
2-Fluorophenol	0.932	0.835		-10.4	50.0
Phenol-d6	1.152	0.978		-15.1	50.0
Nitrobenzene-d5	0.435	0.364		-16.3	50.0
2-Fluorobiphenyl	2.327	1.915		-17.7	50.0
2,4,6-Tribromophenol	0.182	0.169		-7.1	50.0
Terphenyl-d14	0.958	0.875		-8.7	50.0
1,4-Dioxane	0.444	0.456		2.7	50.0

All other compounds must meet a minimum RRF of 0.010.