

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032825\
 Data File : BN036791.D
 Acq On : 29 Mar 2025 02:44
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 29 04:28:30 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 10 16:06:28 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	-0.03
2	1,4-Dioxane	0.444	0.453	-2.0	75	-0.01
3	n-Nitrosodimethylamine	0.898	0.962	-7.1	85	0.00
4 S	2-Fluorophenol	0.932	0.833	10.6	70	-0.02
5 S	Phenol-d6	1.152	1.007	12.6	74	-0.03
6	bis(2-Chloroethyl)ether	1.190	1.030	13.4	72	-0.03
7 I	Naphthalene-d8	1.000	1.000	0.0	88	-0.03
8 S	Nitrobenzene-d5	0.435	0.376	13.6	80	-0.03
9	Naphthalene	1.177	1.056	10.3	77	-0.03
10	Hexachlorobutadiene	0.277	0.258	6.9	77	-0.03
11 SURR	2-Methylnaphthalene-d10	0.595	0.526	11.6	77	-0.04
12	2-Methylnaphthalene	0.749	0.678	9.5	78	-0.03
13 I	Acenaphthene-d10	1.000	1.000	0.0	89	-0.03
14 S	2,4,6-Tribromophenol	0.182	0.178	2.2	85	-0.02
15 S	2-Fluorobiphenyl	2.327	2.041	12.3	76	-0.03
16	Acenaphthylene	1.888	1.678	11.1	77	-0.02
17	Acenaphthene	1.236	1.134	8.3	79	-0.03
18	Fluorene	1.672	1.558	6.8	79	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	93	-0.02
20	4,6-Dinitro-2-methylphenol	0.086	0.071	17.4	85	-0.02
21	4-Bromophenyl-phenylether	0.251	0.243	3.2	82	-0.02
22	Hexachlorobenzene	0.303	0.284	6.3	78	-0.02
23	Atrazine	0.201	0.196	2.5	85	-0.02
24	Pentachlorophenol	0.138	0.133	3.6	90	-0.04
25	Phenanthrene	1.200	1.152	4.0	82	-0.02
26	Anthracene	1.083	1.018	6.0	82	-0.02
27 SURR	Fluoranthene-d10	1.025	1.032	-0.7	86	-0.02
28	Fluoranthene	1.348	1.397	-3.6	89	-0.03
29 I	Chrysene-d12	1.000	1.000	0.0	113	-0.03
30	Pyrene	1.956	1.715	12.3	91	-0.02
31 S	Terphenyl-d14	0.958	0.782	18.4	86	-0.02
32	Benzo(a)anthracene	1.391	1.261	9.3	99	-0.03
33	Chrysene	1.520	1.463	3.7	102	-0.03
34	Bis(2-ethylhexyl)phthalate	0.990	0.940	5.1	101	-0.02
35 I	Perylene-d12	1.000	1.000	0.0	114	-0.04
36	Indeno(1,2,3-cd)pyrene	1.444	1.341	7.1	99	-0.06
37	Benzo(b)fluoranthene	1.456	1.396	4.1	103	-0.03
38	Benzo(k)fluoranthene	1.527	1.527	0.0	107	-0.03
39 C	Benzo(a)pyrene	1.226	1.238	-1.0	108	-0.04
40	Dibenzo(a,h)anthracene	1.124	1.029	8.5	101	-0.06
41	Benzo(g,h,i)perylene	1.286	1.209	6.0	100	-0.06

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4