

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN032825\
 Data File : BN036807.D
 Acq On : 29 Mar 2025 12:21
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 31 02:01:08 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 10 16:06:28 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	-0.03
2	1,4-Dioxane	0.444	0.456	-2.7	82	-0.01
3	n-Nitrosodimethylamine	0.898	0.931	-3.7	89	-0.01
4 S	2-Fluorophenol	0.932	0.835	10.4	76	-0.02
5 S	Phenol-d6	1.152	0.978	15.1	78	-0.04
6	bis(2-Chloroethyl)ether	1.190	1.007	15.4	76	-0.03
7 I	Naphthalene-d8	1.000	1.000	0.0	97	-0.03
8 S	Nitrobenzene-d5	0.435	0.364	16.3	85	-0.03
9	Naphthalene	1.177	1.046	11.1	84	-0.03
10	Hexachlorobutadiene	0.277	0.248	10.5	82	-0.03
11 SURR	2-Methylnaphthalene-d10	0.595	0.526	11.6	85	-0.04
12	2-Methylnaphthalene	0.749	0.668	10.8	85	-0.03
13 I	Acenaphthene-d10	1.000	1.000	0.0	101	-0.03
14 S	2,4,6-Tribromophenol	0.182	0.169	7.1	90	-0.02
15 S	2-Fluorobiphenyl	2.327	1.915	17.7	80	-0.03
16	Acenaphthylene	1.888	1.643	13.0	85	-0.02
17	Acenaphthene	1.236	1.115	9.8	88	-0.03
18	Fluorene	1.672	1.519	9.2	87	-0.03
19 I	Phenanthrene-d10	1.000	1.000	0.0	103	-0.02
20	4,6-Dinitro-2-methylphenol	0.086	0.077	10.5	103	-0.02
21	4-Bromophenyl-phenylether	0.251	0.231	8.0	87	-0.02
22	Hexachlorobenzene	0.303	0.280	7.6	86	-0.02
23	Atrazine	0.201	0.189	6.0	91	-0.02
24	Pentachlorophenol	0.138	0.128	7.2	96	-0.04
25	Phenanthrene	1.200	1.132	5.7	90	-0.02
26	Anthracene	1.083	1.013	6.5	91	-0.04
27 SURR	Fluoranthene-d10	1.025	0.985	3.9	91	-0.02
28	Fluoranthene	1.348	1.322	1.9	94	-0.03
29 I	Chrysene-d12	1.000	1.000	0.0	106	-0.03
30	Pyrene	1.956	1.894	3.2	94	-0.03
31 S	Terphenyl-d14	0.958	0.875	8.7	90	-0.02
32	Benzo(a)anthracene	1.391	1.232	11.4	91	-0.03
33	Chrysene	1.520	1.434	5.7	95	-0.03
34	Bis(2-ethylhexyl)phthalate	0.990	0.953	3.7	97	-0.02
35 I	Perylene-d12	1.000	1.000	0.0	108	-0.04
36	Indeno(1,2,3-cd)pyrene	1.444	1.359	5.9	95	-0.06
37	Benzo(b)fluoranthene	1.456	1.360	6.6	95	-0.03
38	Benzo(k)fluoranthene	1.527	1.528	-0.1	102	-0.03
39 C	Benzo(a)pyrene	1.226	1.191	2.9	99	-0.04
40	Dibenzo(a,h)anthracene	1.124	1.077	4.2	100	-0.06
41	Benzo(g,h,i)perylene	1.286	1.194	7.2	93	-0.06

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4