DATA OF KNOWN QUALITY CONFORMANCE/NON-CONFORMANCE SUMMARY QUESTIONNAIRE

Labora	atory Name : Alliance Technical Group LLC	Client :	G Environmental					
Project Location : Project Number : - Buffington - GEO			ECP					
Laboratory Sample ID(s): Q1622 Sampling Date(s): 3/19/2025								
List Dh	List DKQP Methods Used (e.g., 8260,8270, et Cetra) 8260D,SOP							
1	For each analytical method referenced in this laboratory respecified QA/QC performance criteria followed, including the explain any criteria falling outside of acceptable guidelines NJDEP Data of Known Quality performance standards?	the requirement to		\	Yes		No	
1A	Were the method specified handling, preservation, and ho	olding time requirements	s met?	Ø	Yes		No	
1B	EPH Method: Was the EPH method conducted without sig (see Section 11.3 of respective DKQ methods)	gnificant modifications			Yes		No	☑ N/A
2	Were all samples received by the laboratory in a condition described on the associated chain-of-custody document(s			V	Yes		No	
3	Were samples received at an appropriate temperature (4	±2° C)?		V	Yes		No	□ N/A
4	Were all QA/QC performance criteria specified in the NJD standards achieved?	EP DKQP			Yes	V	No	
5	a)Were reporting limits specified or referenced on the cha communicated to the laboratory prior to sample receipt?	in-of-custody or		V	Yes		No	
	b)Were these reporting limits met?			V	Yes		No	□ N/A
6	For each analytical method referenced in this laboratory results reported for all constituents identified in the method presented in the DKQP documents and/or site-specific QA	od-specific analyte lists		V	Yes		No	
7	Are project-specific matrix spikes and/or laboratory duplication	ates included in this dat	a set?	V	Yes		No	

Notes: For all questions to which the response was "No" (with the exception of question #7), additional information should be provided in an attached narrative. If the answer to question #1, #1A, or #1B is "No", the data package does not meet the requirements for "Data of Known Quality."



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Cover Page

Order ID:	Q1622
-----------	-------

Project ID: Buffington - GECP

Client: G Environmental

Lab Sample Number Client Sample Number

Q1622-01 MW1
Q1622-02 MW3
Q1622-03 Field-Blank

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature :		
Signature .	——————————————————————————————————————	· 3/27/202

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012