

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN031825

Review By	John Carlone	Review On	3/19/2025 9:43:20 AM
Supervise By	Maresh Dadoda	Supervise On	3/19/2025 2:22:13 PM
SubDirectory	VN031825	HP Acquire Method	HP Processing Method 82N031825W.M

STD. NAME	STD REF.#
Tune/Reschk	VP133352
Initial Calibration Stds	VP133393,VP133394,VP133395,VP133396,VP133397,VP133399
CCC	VP133353
Internal Standard/PEM	
ICV/I.BLK	VP133401
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN085994.D	18 Mar 2025 08:52		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN085995.D	18 Mar 2025 10:55	Need ICAL	JC\MD	Not Ok
3	VSTDICC100	VSTDICC100	VN085996.D	18 Mar 2025 11:44		JC\MD	Ok,M
4	VSTDICCC050	VSTDICCC050	VN085997.D	18 Mar 2025 12:08	Comp.#56 and 58 is on Quadratic Regression	JC\MD	Ok,M
5	VSTDICC020	VSTDICC020	VN085998.D	18 Mar 2025 12:32		JC\MD	Ok,M
6	VSTDICC010	VSTDICC010	VN085999.D	18 Mar 2025 12:57	%D failed for Comp. #58 in 01PPB and 20PPB	JC\MD	Ok,M
7	VSTDICC005	VSTDICC005	VN086000.D	18 Mar 2025 13:21		JC\MD	Ok,M
8	VSTDICC001	VSTDICC001	VN086001.D	18 Mar 2025 14:09		JC\MD	Ok,M
9	IBLK	IBLK	VN086002.D	18 Mar 2025 14:34		JC\MD	Ok
10	VSTDICV050	ICVVN031825	VN086003.D	18 Mar 2025 16:49	ICV Failed for comp. #39,52,58,68,69,70,78,80,84,85, 86,89 For DOD	JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN032425

Review By	John Carlone	Review On	3/25/2025 9:56:11 AM
Supervise By	Maresh Dadoda	Supervise On	3/25/2025 3:26:57 PM
SubDirectory	VN032425	HP Acquire Method	HP Processing Method 82N031825W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133454
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133455,VP133456

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN086064.D	24 Mar 2025 11:38		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN086065.D	24 Mar 2025 12:03	pH#Lot#V12668	JC\MD	Ok,M
3	VN0324MBL01	VN0324MBL01	VN086066.D	24 Mar 2025 12:38		JC\MD	Ok
4	VN0324WBL01	VN0324WBL01	VN086067.D	24 Mar 2025 13:02		JC\MD	Ok
5	VN0324MBS01	VN0324MBS01	VN086068.D	24 Mar 2025 13:26	MBS Failed low for com.#93	JC\MD	Ok,M
6	VN0324WBS01	VN0324WBS01	VN086069.D	24 Mar 2025 14:00		JC\MD	Ok,M
7	Q1598-01ME	ETGI-354ME	VN086070.D	24 Mar 2025 14:24	Need 500X	JC\MD	Dilution
8	IBLK	IBLK	VN086071.D	24 Mar 2025 14:48		JC\MD	Ok
9	IBLK	IBLK	VN086072.D	24 Mar 2025 15:12		JC\MD	Ok
10	Q1598-01ME	ETGI-354ME	VN086073.D	24 Mar 2025 15:36	Need further dilution	JC\MD	Not Ok
11	Q1606-01	CHRT24743	VN086074.D	24 Mar 2025 16:00		JC\MD	Ok
12	Q1606-07	RT3025	VN086075.D	24 Mar 2025 16:25		JC\MD	Ok
13	Q1606-09	CHRT28607	VN086076.D	24 Mar 2025 16:49		JC\MD	Ok
14	VN0324WBSD01	VN0324WBSD01	VN086077.D	24 Mar 2025 17:13		JC\MD	Ok,M
15	Q1598-01MEDL	ETGI-354MEDL	VN086078.D	24 Mar 2025 17:37		JC\MD	Ok
16	Q1619-08	TP-2	VN086079.D	24 Mar 2025 18:01	vial A pH#5.0 Internal standard fail	JC\MD	ReRun
17	Q1619-10	TP-3	VN086080.D	24 Mar 2025 18:25	vial A pH#5.0 Internal standard fail	JC\MD	ReRun

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN032425

Review By	John Carlone	Review On	3/25/2025 9:56:11 AM
Supervise By	Maresh Dadoda	Supervise On	3/25/2025 3:26:57 PM
SubDirectory	VN032425	HP Acquire Method	HP Processing Method 82N031825W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133454		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133455,VP133456		

18	Q1619-12	TP-4	VN086081.D	24 Mar 2025 18:49	vial A pH#5.0 Internal standard fail	JC\MD	ReRun
19	Q1619-14	TP-5	VN086082.D	24 Mar 2025 19:13	vial A pH#5.0 Internal standard fail	JC\MD	ReRun
20	Q1625-01	FRAC-TANK-A1291	VN086083.D	24 Mar 2025 19:37	vial A pH<2	JC\MD	Ok
21	Q1618-02	TP20250320-02	VN086084.D	24 Mar 2025 20:01	Need Straight Run- run 10x	JC\MD	Not Ok
22	Q1622-03	Field-Blank	VN086085.D	24 Mar 2025 20:26	vial A pH<2 FB	JC\MD	Ok
23	Q1622-01	MW1	VN086086.D	24 Mar 2025 20:50	vial A pH<2	JC\MD	Ok
24	Q1622-02	MW3	VN086087.D	24 Mar 2025 21:14	vial A pH<2	JC\MD	Ok
25	VSTDCCC050	VSTDCCC050EC	VN086088.D	24 Mar 2025 21:38		JC\MD	Ok,M

M : Manual Integration