

LAB CHRONICLE

OrderID:	Q1622	OrderDate:	3/21/2025 10:43:00 AM
Client:	G Environmental	Project:	Buffington - GECP
Contact:	Gary Landis	Location:	VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1622-01	MW1	Water	VOC-TCLVOA-10	8260-Low	03/19/25		03/24/25	03/21/25
Q1622-02	MW3	Water	VOC-TCLVOA-10	8260-Low	03/19/25		03/24/25	03/21/25
Q1622-03	Field-Blank	Water	VOC-TCLVOA-10	8260-Low	03/19/25		03/24/25	03/21/25

Hit Summary Sheet
SW-846

SDG No.: Q1622
Client: G Environmental

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	MW1							
Q1622-01	MW1	Water	Acetone	2.60	J	1.50	5.00	ug/L
Q1622-01	MW1	Water	Chloroform	0.81	J	0.25	1.00	ug/L
Q1622-01	MW1	Water	Trichloroethene	0.31	J	0.090	1.00	ug/L
Q1622-01	MW1	Water	Tetrachloroethene	0.54	J	0.23	1.00	ug/L
			Total Voc :	4.26				
			Total Concentration:	4.26				
Client ID:	MW3							
Q1622-02	MW3	Water	Acetone	3.30	J	1.50	5.00	ug/L
			Total Voc :	3.30				
			Total Concentration:	3.30				



SAMPLE DATA

Report of Analysis

Client:	G Environmental	Date Collected:	03/19/25
Project:	Buffington - GECP	Date Received:	03/21/25
Client Sample ID:	MW1	SDG No.:	Q1622
Lab Sample ID:	Q1622-01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086086.D	1		03/24/25 20:50	VN032425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	2.60	J	1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	U	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	0.81	J	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.31	J	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	U	0.68	5.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	G Environmental	Date Collected:	03/19/25
Project:	Buffington - GECP	Date Received:	03/21/25
Client Sample ID:	MW1	SDG No.:	Q1622
Lab Sample ID:	Q1622-01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086086.D	1		03/24/25 20:50	VN032425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.17	U	0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.16	U	0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
591-78-6	2-Hexanone	0.89	U	0.89	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.15	U	0.15	1.00	ug/L
127-18-4	Tetrachloroethene	0.54	J	0.23	1.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	1.00	ug/L
100-41-4	Ethyl Benzene	0.13	U	0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	0.24	U	0.24	2.00	ug/L
95-47-6	o-Xylene	0.12	UQ	0.12	1.00	ug/L
100-42-5	Styrene	0.15	U	0.15	1.00	ug/L
75-25-2	Bromoform	0.19	U	0.19	1.00	ug/L
98-82-8	Isopropylbenzene	0.12	UQ	0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	0.26	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.6		74 - 125	103%	SPK: 50
1868-53-7	Dibromofluoromethane	49.1		75 - 124	98%	SPK: 50
2037-26-5	Toluene-d8	46.2		86 - 113	92%	SPK: 50
460-00-4	4-Bromofluorobenzene	42.6		77 - 121	85%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	253000	8.224			
540-36-3	1,4-Difluorobenzene	462000	9.1			
3114-55-4	Chlorobenzene-d5	399000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	163000	13.788			

Report of Analysis

Client:	G Environmental	Date Collected:	03/19/25
Project:	Buffington - GECP	Date Received:	03/21/25
Client Sample ID:	MW1	SDG No.:	Q1622
Lab Sample ID:	Q1622-01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	RXI-624	ID :	0.25
Prep Method :		Level :	LOW
		Final Vol:	5000
		Test:	VOC-TCLVOA-10

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086086.D	1		03/24/25 20:50	VN032425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	G Environmental	Date Collected:	03/19/25
Project:	Buffington - GECP	Date Received:	03/21/25
Client Sample ID:	MW3	SDG No.:	Q1622
Lab Sample ID:	Q1622-02	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086087.D	1		03/24/25 21:14	VN032425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	3.30	J	1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	U	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	U	0.68	5.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	G Environmental		Date Collected:	03/19/25	
Project:	Buffington - GECP		Date Received:	03/21/25	
Client Sample ID:	MW3		SDG No.:	Q1622	
Lab Sample ID:	Q1622-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086087.D	1		03/24/25 21:14	VN032425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.17	U	0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.16	U	0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
591-78-6	2-Hexanone	0.89	U	0.89	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.15	U	0.15	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	1.00	ug/L
100-41-4	Ethyl Benzene	0.13	U	0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	0.24	U	0.24	2.00	ug/L
95-47-6	o-Xylene	0.12	UQ	0.12	1.00	ug/L
100-42-5	Styrene	0.15	U	0.15	1.00	ug/L
75-25-2	Bromoform	0.19	U	0.19	1.00	ug/L
98-82-8	Isopropylbenzene	0.12	UQ	0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	0.26	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.1		74 - 125	104%	SPK: 50
1868-53-7	Dibromofluoromethane	48.7		75 - 124	97%	SPK: 50
2037-26-5	Toluene-d8	46.5		86 - 113	93%	SPK: 50
460-00-4	4-Bromofluorobenzene	43.4		77 - 121	87%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	243000	8.224			
540-36-3	1,4-Difluorobenzene	452000	9.1			
3114-55-4	Chlorobenzene-d5	399000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	157000	13.788			

Report of Analysis

Client:	G Environmental	Date Collected:	03/19/25
Project:	Buffington - GECP	Date Received:	03/21/25
Client Sample ID:	MW3	SDG No.:	Q1622
Lab Sample ID:	Q1622-02	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5	Units:	mL
Soil Aliquot Vol:			uL
GC Column:	RXI-624	ID :	0.25
Prep Method :		Level :	LOW

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086087.D	1		03/24/25 21:14	VN032425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	G Environmental	Date Collected:	03/19/25
Project:	Buffington - GECP	Date Received:	03/21/25
Client Sample ID:	Field-Blank	SDG No.:	Q1622
Lab Sample ID:	Q1622-03	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086085.D	1		03/24/25 20:26	VN032425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	1.50	U	1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	U	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	U	0.68	5.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	G Environmental		Date Collected:	03/19/25	
Project:	Buffington - GECP		Date Received:	03/21/25	
Client Sample ID:	Field-Blank		SDG No.:	Q1622	
Lab Sample ID:	Q1622-03		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086085.D	1		03/24/25 20:26	VN032425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.17	U	0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.16	U	0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
591-78-6	2-Hexanone	0.89	U	0.89	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.15	U	0.15	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	1.00	ug/L
100-41-4	Ethyl Benzene	0.13	U	0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	0.24	U	0.24	2.00	ug/L
95-47-6	o-Xylene	0.12	UQ	0.12	1.00	ug/L
100-42-5	Styrene	0.15	U	0.15	1.00	ug/L
75-25-2	Bromoform	0.19	U	0.19	1.00	ug/L
98-82-8	Isopropylbenzene	0.12	UQ	0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	0.26	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.5		74 - 125	103%	SPK: 50
1868-53-7	Dibromofluoromethane	48.5		75 - 124	97%	SPK: 50
2037-26-5	Toluene-d8	45.9		86 - 113	92%	SPK: 50
460-00-4	4-Bromofluorobenzene	43.0		77 - 121	86%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	262000	8.224			
540-36-3	1,4-Difluorobenzene	484000	9.1			
3114-55-4	Chlorobenzene-d5	422000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	173000	13.788			



QC SUMMARY

Surrogate Summary

SDG No.: **Q1622**

Client: **G Environmental**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1622-01	MW1	1,2-Dichloroethane-d4	50	51.6	103	74	125
		Dibromofluoromethane	50	49.1	98	75	124
		Toluene-d8	50	46.2	92	86	113
		4-Bromofluorobenzene	50	42.6	85	77	121
Q1622-02	MW3	1,2-Dichloroethane-d4	50	52.1	104	74	125
		Dibromofluoromethane	50	48.7	97	75	124
		Toluene-d8	50	46.5	93	86	113
		4-Bromofluorobenzene	50	43.4	87	77	121
Q1622-03	Field-Blank	1,2-Dichloroethane-d4	50	51.5	103	74	125
		Dibromofluoromethane	50	48.5	97	75	124
		Toluene-d8	50	45.9	92	86	113
		4-Bromofluorobenzene	50	43.0	86	77	121
VN0324WBL01	VN0324WBL01	1,2-Dichloroethane-d4	50	57.9	116	74	125
		Dibromofluoromethane	50	56.0	112	75	124
		Toluene-d8	50	49.1	98	86	113
		4-Bromofluorobenzene	50	42.4	85	77	121
VN0324WBS01	VN0324WBS01	1,2-Dichloroethane-d4	50	51.0	102	74	125
		Dibromofluoromethane	50	50.8	102	75	124
		Toluene-d8	50	50.8	102	86	113
		4-Bromofluorobenzene	50	50.4	101	77	121
VN0324WBSD0	VN0324WBSD01	1,2-Dichloroethane-d4	50	46.4	93	74	125
		Dibromofluoromethane	50	44.9	90	75	124
		Toluene-d8	50	49.8	100	86	113
		4-Bromofluorobenzene	50	52.2	104	77	121

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1622

Client: G Environmental

Analytical Method: SW8260-Low

Datafile : VN086069.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VN0324WBS01	Dichlorodifluoromethane	20	18.3	ug/L	92			69	116	
	Chloromethane	20	18.1	ug/L	91			65	116	
	Vinyl chloride	20	18.7	ug/L	94			65	117	
	Bromomethane	20	17.9	ug/L	90			58	125	
	Chloroethane	20	17.8	ug/L	89			56	128	
	Trichlorofluoromethane	20	17.5	ug/L	88			73	115	
	1,1,2-Trichlorotrifluoroethane	20	18.6	ug/L	93			80	112	
	1,1-Dichloroethene	20	17.9	ug/L	90			74	110	
	Acetone	100	85.3	ug/L	85			60	125	
	Carbon disulfide	20	16.4	ug/L	82			64	112	
	Methyl tert-butyl Ether	20	17.8	ug/L	89			78	114	
	Methyl Acetate	20	20.2	ug/L	101			67	125	
	Methylene Chloride	20	18.9	ug/L	95			72	114	
	trans-1,2-Dichloroethene	20	18.2	ug/L	91			75	108	
	1,1-Dichloroethane	20	19.0	ug/L	95			78	112	
	Cyclohexane	20	17.1	ug/L	86			75	110	
	2-Butanone	100	87.2	ug/L	87			65	122	
	Carbon Tetrachloride	20	18.1	ug/L	91			77	113	
	cis-1,2-Dichloroethene	20	18.5	ug/L	93			77	110	
	Bromochloromethane	20	20.2	ug/L	101			70	124	
	Chloroform	20	18.8	ug/L	94			79	113	
	1,1,1-Trichloroethane	20	17.9	ug/L	90			80	108	
	Methylcyclohexane	20	16.6	ug/L	83			72	115	
	Benzene	20	18.9	ug/L	95			82	109	
	1,2-Dichloroethane	20	18.9	ug/L	95			80	115	
	Trichloroethene	20	16.7	ug/L	84			77	113	
	1,2-Dichloropropane	20	19.6	ug/L	98			83	111	
	Bromodichloromethane	20	19.0	ug/L	95			83	110	
	4-Methyl-2-Pentanone	100	96.0	ug/L	96			74	118	
	Toluene	20	19.9	ug/L	100			82	110	
	t-1,3-Dichloropropene	20	18.8	ug/L	94			79	110	
	cis-1,3-Dichloropropene	20	18.4	ug/L	92			82	110	
	1,1,2-Trichloroethane	20	20.0	ug/L	100			83	112	
	2-Hexanone	100	94.7	ug/L	95			73	117	
	Dibromochloromethane	20	19.5	ug/L	98			82	110	
	1,2-Dibromoethane	20	19.3	ug/L	97			81	110	
	Tetrachloroethene	20	18.5	ug/L	93			67	123	
	Chlorobenzene	20	18.1	ug/L	91			82	109	
	Ethyl Benzene	20	17.8	ug/L	89			83	109	
	m/p-Xylenes	40	36.9	ug/L	92			82	110	
	o-Xylene	20	18.4	ug/L	92			83	109	
	Styrene	20	18.3	ug/L	92			80	111	
	Bromoform	20	18.4	ug/L	92			79	109	
	Isopropylbenzene	20	16.9	ug/L	85			83	112	
	1,1,2,2-Tetrachloroethane	20	17.3	ug/L	86			76	118	
	1,3-Dichlorobenzene	20	17.9	ug/L	90			82	108	
	1,4-Dichlorobenzene	20	17.0	ug/L	85			82	107	
	1,2-Dichlorobenzene	20	16.9	ug/L	85			82	109	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1622

Client: G Environmental

Analytical Method: SW8260-Low

Datafile : VN086069.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VN0324WBS01	1,2-Dibromo-3-Chloropropane	20	17.0	ug/L	85			68	112	
	1,2,4-Trichlorobenzene	20	15.6	ug/L	78			75	113	
	1,2,3-Trichlorobenzene	20	15.4	ug/L	77			76	114	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1622

Client: G Environmental

Analytical Method: SW8260-Low

Datafile : VN086077.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits			
									High	RPD		
VN0324WBSD01	Dichlorodifluoromethane	20	15.9	ug/L	79	15		69	116	20		
	Chloromethane	20	18.3	ug/L	92	1		65	116	20		
	Vinyl chloride	20	19.5	ug/L	98	4		65	117	20		
	Bromomethane	20	17.4	ug/L	87	3		58	125	20		
	Chloroethane	20	19.0	ug/L	95	7		56	128	20		
	Trichlorofluoromethane	20	16.7	ug/L	84	5		73	115	20		
	1,1,2-Trichlorotrifluoroethane	20	18.5	ug/L	93	0		80	112	20		
	1,1-Dichloroethene	20	19.0	ug/L	95	5		74	110	20		
	Acetone	100	83.5	ug/L	84	1		60	125	20		
	Carbon disulfide	20	16.1	ug/L	81	1		64	112	20		
	Methyl tert-butyl Ether	20	19.4	ug/L	97	9		78	114	20		
	Methyl Acetate	20	21.5	ug/L	108	7		67	125	20		
	Methylene Chloride	20	18.9	ug/L	95	0		72	114	20		
	trans-1,2-Dichloroethene	20	19.3	ug/L	97	6		75	108	20		
	1,1-Dichloroethane	20	19.6	ug/L	98	3		78	112	20		
	Cyclohexane	20	20.3	ug/L	102	17		75	110	20		
	2-Butanone	100	95.6	ug/L	96	10		65	122	20		
	Carbon Tetrachloride	20	16.8	ug/L	84	8		77	113	20		
	cis-1,2-Dichloroethene	20	19.9	ug/L	100	7		77	110	20		
	Bromochloromethane	20	20.9	ug/L	104	3		70	124	20		
	Chloroform	20	19.0	ug/L	95	1		79	113	20		
	1,1,1-Trichloroethane	20	18.8	ug/L	94	4		80	108	20		
	Methylcyclohexane	20	21.4	ug/L	107	25		*	72	115	20	
	Benzene	20	19.0	ug/L	95	0		82	109	20		
	1,2-Dichloroethane	20	16.9	ug/L	85	11		80	115	20		
	Trichloroethene	20	18.2	ug/L	91	8		77	113	20		
	1,2-Dichloropropane	20	20.6	ug/L	103	5		83	111	20		
	Bromodichloromethane	20	18.3	ug/L	92	3		83	110	20		
	4-Methyl-2-Pentanone	100	100	ug/L	100	4		74	118	20		
	Toluene	20	20.5	ug/L	103	3		82	110	20		
	t-1,3-Dichloropropene	20	19.6	ug/L	98	4		79	110	20		
	cis-1,3-Dichloropropene	20	19.8	ug/L	99	7		82	110	20		
	1,1,2-Trichloroethane	20	20.8	ug/L	104	4		83	112	20		
	2-Hexanone	100	110	ug/L	110	15		73	117	20		
	Dibromochloromethane	20	18.0	ug/L	90	9		82	110	20		
	1,2-Dibromoethane	20	19.8	ug/L	99	2		81	110	20		
	Tetrachloroethene	20	18.1	ug/L	91	2		67	123	20		
	Chlorobenzene	20	18.6	ug/L	93	2		82	109	20		
	Ethyl Benzene	20	20.5	ug/L	103	15		83	109	20		
	m/p-Xylenes	40	40.1	ug/L	100	8		82	110	20		
	o-Xylene	20	21.9	ug/L	110	18		*	83	109	20	
	Styrene	20	20.9	ug/L	104	12			80	111	20	
	Bromoform	20	17.5	ug/L	88	4			79	109	20	
	Isopropylbenzene	20	22.5	ug/L	113	28		*	*	83	112	20
	1,1,2,2-Tetrachloroethane	20	18.6	ug/L	93	8			76	118	20	
	1,3-Dichlorobenzene	20	19.2	ug/L	96	6			82	108	20	
	1,4-Dichlorobenzene	20	17.9	ug/L	90	6			82	107	20	
	1,2-Dichlorobenzene	20	19.2	ug/L	96	12			82	109	20	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1622

Client: G Environmental

Analytical Method: SW8260-Low

Datafile : VN086077.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0324WBSD01	1,2-Dibromo-3-Chloropropane	20	18.8	ug/L	94	10		68	112	20
	1,2,4-Trichlorobenzene	20	20.3	ug/L	102	27	*	75	113	20
	1,2,3-Trichlorobenzene	20	21.3	ug/L	106	32	*	76	114	20

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VN0324WBL01

Lab Name: CHEMTECH

Contract: GENV01

Lab Code: CHEM Case No.: Q1622

SAS No.: Q1622 SDG NO.: Q1622

Lab File ID: VN086067.D

Lab Sample ID: VN0324WBL01

Date Analyzed: 03/24/2025

Time Analyzed: 13:02

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_N

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VN0324WBS01	VN0324WBS01	VN086069.D	03/24/2025
VN0324WBSD01	VN0324WBSD01	VN086077.D	03/24/2025
Field-Blank	Q1622-03	VN086085.D	03/24/2025
MW1	Q1622-01	VN086086.D	03/24/2025
MW3	Q1622-02	VN086087.D	03/24/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q1622 SAS No.: Q1622 SDG NO.: Q1622
Lab File ID: VN085994.D BFB Injection Date: 03/18/2025
Instrument ID: MSVOA_N BFB Injection Time: 08:52
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	15.2
75	30.0 - 60.0% of mass 95	52.9
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.9
173	Less than 2.0% of mass 174	0.6 (0.7) 1
174	50.0 - 100.0% of mass 95	87.1
175	5.0 - 9.0% of mass 174	6.7 (7.7) 1
176	95.0 - 101.0% of mass 174	83.7 (96.2) 1
177	5.0 - 9.0% of mass 176	5.4 (6.4) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICC100	VSTDICC100	VN085996.D	03/18/2025	11:44
VSTDICCC050	VSTDICCC050	VN085997.D	03/18/2025	12:08
VSTDICC020	VSTDICC020	VN085998.D	03/18/2025	12:32
VSTDICC010	VSTDICC010	VN085999.D	03/18/2025	12:57
VSTDICC005	VSTDICC005	VN086000.D	03/18/2025	13:21
VSTDICC001	VSTDICC001	VN086001.D	03/18/2025	14:09

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q1622 SAS No.: Q1622 SDG NO.: Q1622
Lab File ID: VN086064.D BFB Injection Date: 03/24/2025
Instrument ID: MSVOA_N BFB Injection Time: 11:38
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	16.4
75	30.0 - 60.0% of mass 95	51.3
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.3
173	Less than 2.0% of mass 174	1.6 (1.8) 1
174	50.0 - 100.0% of mass 95	89.6
175	5.0 - 9.0% of mass 174	6.5 (7.2) 1
176	95.0 - 101.0% of mass 174	86 (96) 1
177	5.0 - 9.0% of mass 176	6.1 (7.1) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN086065.D	03/24/2025	12:03
VN0324WBL01	VN0324WBL01	VN086067.D	03/24/2025	13:02
VN0324WBS01	VN0324WBS01	VN086069.D	03/24/2025	14:00
VN0324WBSD01	VN0324WBSD01	VN086077.D	03/24/2025	17:13
Field-Blank	Q1622-03	VN086085.D	03/24/2025	20:26
MW1	Q1622-01	VN086086.D	03/24/2025	20:50
MW3	Q1622-02	VN086087.D	03/24/2025	21:14

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q1622 SAS No.: Q1622 SDG NO.: Q1622
Lab File ID: VN086065.D Date Analyzed: 03/24/2025
Instrument ID: MSVOA_N Time Analyzed: 12:03
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	160831	8.22	254989	9.10	239260	11.87
UPPER LIMIT	321662	8.724	509978	9.6	478520	12.365
LOWER LIMIT	80415.5	7.724	127495	8.6	119630	11.365
EPA SAMPLE NO.						
MW1	252685	8.22	462481	9.10	398785	11.87
MW3	243235	8.22	452406	9.10	398702	11.87
Field-Blank	262167	8.22	483807	9.10	422031	11.87
VN0324WBL01	131247	8.22	230937	9.10	213386	11.87
VN0324WBS01	162414	8.22	258911	9.10	235727	11.87
VN0324WBSD01	270759	8.22	461280	9.10	433499	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1622 SAS No.: Q1622 SDG NO.: Q1622
 Lab File ID: VN086065.D Date Analyzed: 03/24/2025
 Instrument ID: MSVOA_N Time Analyzed: 12:03
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	132834	13.788				
UPPER LIMIT	265668	14.288				
LOWER LIMIT	66417	13.288				
EPA SAMPLE NO.						
MW1	163397	13.79				
MW3	157433	13.79				
Field-Blank	173343	13.79				
VN0324WBL01	85734	13.79				
VN0324WBS01	121424	13.79				
VN0324WBSD01	206640	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	G Environmental		Date Collected:	
Project:	Buffington - GECP		Date Received:	
Client Sample ID:	VN0324WBL01		SDG No.:	Q1622
Lab Sample ID:	VN0324WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086067.D	1		03/24/25 13:02	VN032425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	1.50	U	1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	U	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	U	0.68	5.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	G Environmental		Date Collected:	
Project:	Buffington - GECP		Date Received:	
Client Sample ID:	VN0324WBL01		SDG No.:	Q1622
Lab Sample ID:	VN0324WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086067.D	1		03/24/25 13:02	VN032425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.17	U	0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.16	U	0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
591-78-6	2-Hexanone	0.89	U	0.89	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.15	U	0.15	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	1.00	ug/L
100-41-4	Ethyl Benzene	0.13	U	0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	0.24	U	0.24	2.00	ug/L
95-47-6	o-Xylene	0.12	U	0.12	1.00	ug/L
100-42-5	Styrene	0.15	U	0.15	1.00	ug/L
75-25-2	Bromoform	0.19	U	0.19	1.00	ug/L
98-82-8	Isopropylbenzene	0.12	U	0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	0.26	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	57.9		74 - 125	116%	SPK: 50
1868-53-7	Dibromofluoromethane	56.0		75 - 124	112%	SPK: 50
2037-26-5	Toluene-d8	49.1		86 - 113	98%	SPK: 50
460-00-4	4-Bromofluorobenzene	42.4		77 - 121	85%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	131000	8.224			
540-36-3	1,4-Difluorobenzene	231000	9.1			
3114-55-4	Chlorobenzene-d5	213000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	85700	13.788			

Report of Analysis

Client:	G Environmental		Date Collected:	
Project:	Buffington - GECP		Date Received:	
Client Sample ID:	VN0324WBL01		SDG No.:	Q1622
Lab Sample ID:	VN0324WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086067.D	1		03/24/25 13:02	VN032425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	G Environmental		Date Collected:	
Project:	Buffington - GECP		Date Received:	
Client Sample ID:	VN0324WBS01		SDG No.:	Q1622
Lab Sample ID:	VN0324WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086069.D	1		03/24/25 14:00	VN032425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	18.3		0.22	1.00	ug/L
74-87-3	Chloromethane	18.1		0.32	1.00	ug/L
75-01-4	Vinyl Chloride	18.7		0.26	1.00	ug/L
74-83-9	Bromomethane	17.9		1.40	5.00	ug/L
75-00-3	Chloroethane	17.8		0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	17.5		0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	18.6		0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	17.9		0.23	1.00	ug/L
67-64-1	Acetone	85.3		1.50	5.00	ug/L
75-15-0	Carbon Disulfide	16.4		0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	17.8		0.16	1.00	ug/L
79-20-9	Methyl Acetate	20.2		0.27	1.00	ug/L
75-09-2	Methylene Chloride	18.9		0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.2		0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.0		0.23	1.00	ug/L
110-82-7	Cyclohexane	17.1		1.50	5.00	ug/L
78-93-3	2-Butanone	87.2		0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	18.1		0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	18.5		0.19	1.00	ug/L
74-97-5	Bromochloromethane	20.2		0.22	1.00	ug/L
67-66-3	Chloroform	18.8		0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	17.9		0.20	1.00	ug/L
108-87-2	Methylcyclohexane	16.6		0.16	1.00	ug/L
71-43-2	Benzene	18.9		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	18.9		0.22	1.00	ug/L
79-01-6	Trichloroethene	16.7		0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	19.6		0.20	1.00	ug/L
75-27-4	Bromodichloromethane	19.0		0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	96.0		0.68	5.00	ug/L
108-88-3	Toluene	19.9		0.14	1.00	ug/L

Report of Analysis

Client:	G Environmental		Date Collected:	
Project:	Buffington - GECP		Date Received:	
Client Sample ID:	VN0324WBS01		SDG No.:	Q1622
Lab Sample ID:	VN0324WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086069.D	1		03/24/25 14:00	VN032425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	18.8		0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	18.4		0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.0		0.21	1.00	ug/L
591-78-6	2-Hexanone	94.7		0.89	5.00	ug/L
124-48-1	Dibromochloromethane	19.5		0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	19.3		0.15	1.00	ug/L
127-18-4	Tetrachloroethene	18.5		0.23	1.00	ug/L
108-90-7	Chlorobenzene	18.1		0.12	1.00	ug/L
100-41-4	Ethyl Benzene	17.8		0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	36.9		0.24	2.00	ug/L
95-47-6	o-Xylene	18.4		0.12	1.00	ug/L
100-42-5	Styrene	18.3		0.15	1.00	ug/L
75-25-2	Bromoform	18.4		0.19	1.00	ug/L
98-82-8	Isopropylbenzene	16.9		0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	17.3		0.26	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	17.9		0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	17.0		0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	16.9		0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	17.0		0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	15.6		0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	15.4		0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.0		74 - 125	102%	SPK: 50
1868-53-7	Dibromofluoromethane	50.8		75 - 124	102%	SPK: 50
2037-26-5	Toluene-d8	50.8		86 - 113	102%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.4		77 - 121	101%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	162000	8.224			
540-36-3	1,4-Difluorobenzene	259000	9.1			
3114-55-4	Chlorobenzene-d5	236000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	121000	13.788			

Report of Analysis

Client:	G Environmental		Date Collected:	
Project:	Buffington - GECP		Date Received:	
Client Sample ID:	VN0324WBS01		SDG No.:	Q1622
Lab Sample ID:	VN0324WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086069.D	1		03/24/25 14:00	VN032425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	G Environmental		Date Collected:	
Project:	Buffington - GECP		Date Received:	
Client Sample ID:	VN0324WBSD01		SDG No.:	Q1622
Lab Sample ID:	VN0324WBSD01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086077.D	1		03/24/25 17:13	VN032425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	15.9		0.22	1.00	ug/L
74-87-3	Chloromethane	18.3		0.32	1.00	ug/L
75-01-4	Vinyl Chloride	19.5		0.26	1.00	ug/L
74-83-9	Bromomethane	17.4		1.40	5.00	ug/L
75-00-3	Chloroethane	19.0		0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	16.7		0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	18.5		0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.0		0.23	1.00	ug/L
67-64-1	Acetone	83.5		1.50	5.00	ug/L
75-15-0	Carbon Disulfide	16.1		0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	19.4		0.16	1.00	ug/L
79-20-9	Methyl Acetate	21.5		0.27	1.00	ug/L
75-09-2	Methylene Chloride	18.9		0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	19.3		0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.6		0.23	1.00	ug/L
110-82-7	Cyclohexane	20.3		1.50	5.00	ug/L
78-93-3	2-Butanone	95.6		0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	16.8		0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.9		0.19	1.00	ug/L
74-97-5	Bromochloromethane	20.9		0.22	1.00	ug/L
67-66-3	Chloroform	19.0		0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	18.8		0.20	1.00	ug/L
108-87-2	Methylcyclohexane	21.4		0.16	1.00	ug/L
71-43-2	Benzene	19.0		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	16.9		0.22	1.00	ug/L
79-01-6	Trichloroethene	18.2		0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	20.6		0.20	1.00	ug/L
75-27-4	Bromodichloromethane	18.3		0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	100		0.68	5.00	ug/L
108-88-3	Toluene	20.5		0.14	1.00	ug/L

Report of Analysis

Client:	G Environmental		Date Collected:	
Project:	Buffington - GECP		Date Received:	
Client Sample ID:	VN0324WBSD01		SDG No.:	Q1622
Lab Sample ID:	VN0324WBSD01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086077.D	1		03/24/25 17:13	VN032425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	19.6		0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	19.8		0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.8		0.21	1.00	ug/L
591-78-6	2-Hexanone	110		0.89	5.00	ug/L
124-48-1	Dibromochloromethane	18.0		0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	19.8		0.15	1.00	ug/L
127-18-4	Tetrachloroethene	18.1		0.23	1.00	ug/L
108-90-7	Chlorobenzene	18.6		0.12	1.00	ug/L
100-41-4	Ethyl Benzene	20.5		0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	40.1		0.24	2.00	ug/L
95-47-6	o-Xylene	21.9		0.12	1.00	ug/L
100-42-5	Styrene	20.9		0.15	1.00	ug/L
75-25-2	Bromoform	17.5		0.19	1.00	ug/L
98-82-8	Isopropylbenzene	22.5		0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	18.6		0.26	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	19.2		0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	17.9		0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	19.2		0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	18.8		0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	20.3		0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	21.3		0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	46.4		74 - 125	93%	SPK: 50
1868-53-7	Dibromofluoromethane	44.9		75 - 124	90%	SPK: 50
2037-26-5	Toluene-d8	49.8		86 - 113	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.2		77 - 121	104%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	271000	8.224			
540-36-3	1,4-Difluorobenzene	461000	9.1			
3114-55-4	Chlorobenzene-d5	433000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	207000	13.788			

Report of Analysis

Client:	G Environmental		Date Collected:	
Project:	Buffington - GECP		Date Received:	
Client Sample ID:	VN0324WBSD01		SDG No.:	Q1622
Lab Sample ID:	VN0324WBSD01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086077.D	1		03/24/25 17:13	VN032425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q1622 SAS No.: Q1622 SDG No.: Q1622

Instrument ID: MSVOA_N Calibration Date(s): 03/18/2025 03/18/2025

Heated Purge: (Y/N) N Calibration Time(s): 11:44 14:09

GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:								
RRF100 = VN085996.D			RRF050 = VN085997.D			RRF020 = VN085998.D		
RRF010 = VN085999.D			RRF005 = VN086000.D			RRF001 = VN086001.D		
COMPOUND	RRF100	RRF050	RRF020	RRF010	RRF005	RRF001	RRF	% RSD
Dichlorodifluoromethane	0.707	0.643	0.600	0.649	0.697	0.678	0.662	6
Chloromethane	0.575	0.557	0.514	0.544	0.612	0.683	0.581	10.3
Vinyl Chloride	0.622	0.592	0.541	0.581	0.634	0.592	0.594	5.5
Bromomethane	0.388	0.381	0.377	0.411	0.466		0.404	9.2
Chloroethane	0.355	0.330	0.334	0.378	0.403	0.446	0.375	11.9
Trichlorofluoromethane	1.059	0.977	0.959	1.035	1.107	1.267	1.067	10.5
1,1,2-Trichlorotrifluoroethane	0.574	0.515	0.512	0.553	0.590	0.659	0.567	9.7
1,1-Dichloroethene	0.567	0.516	0.481	0.545	0.550	0.415	0.512	11
Acetone	0.251	0.194	0.179	0.195	0.213	0.223	0.209	12.4
Carbon Disulfide	1.649	1.510	1.467	1.610	1.765	2.107	1.685	13.8
Methyl tert-butyl Ether	1.888	1.691	1.578	1.604	1.638	1.673	1.679	6.6
Methyl Acetate	0.488	0.469	0.434	0.504	0.542	0.667	0.518	15.8
Methylene Chloride	0.613	0.565	0.551	0.572	0.639	0.731	0.612	10.9
trans-1,2-Dichloroethene	0.603	0.534	0.521	0.546	0.564	0.591	0.560	5.8
1,1-Dichloroethane	1.023	0.949	0.908	0.968	1.032	1.130	1.001	7.8
Cyclohexane	0.890	0.799	0.765	0.854	0.957		0.853	8.9
2-Butanone	0.331	0.292	0.277	0.297	0.319	0.296	0.302	6.5
Carbon Tetrachloride	0.673	0.578	0.557	0.580	0.613	0.624	0.604	6.9
cis-1,2-Dichloroethene	0.688	0.633	0.591	0.617	0.656	0.632	0.636	5.3
Bromochloromethane	0.403	0.391	0.398	0.404	0.430	0.521	0.424	11.6
Chloroform	1.119	1.017	1.011	1.101	1.149	1.245	1.107	7.9
1,1,1-Trichloroethane	1.093	0.986	0.962	1.025	1.097	1.124	1.048	6.3
Methylcyclohexane	0.579	0.475	0.419	0.399	0.426	0.438	0.456	14.3
Benzene	1.610	1.393	1.348	1.386	1.453	1.466	1.443	6.5
1,2-Dichloroethane	0.538	0.475	0.462	0.491	0.528	0.521	0.503	6.1
Trichloroethene	0.394	0.346	0.335	0.353	0.380	0.435	0.374	10
1,2-Dichloropropane	0.366	0.321	0.319	0.321	0.333	0.309	0.328	6.2
Bromodichloromethane	0.600	0.516	0.506	0.515	0.561	0.537	0.539	6.6
4-Methyl-2-Pentanone	0.434	0.385	0.368	0.380	0.369	0.287	0.371	12.9
Toluene	1.074	0.915	0.862	0.878	0.856	0.727	0.885	12.7

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q1622 SAS No.: Q1622 SDG No.: Q1622
Instrument ID: MSVOA_N Calibration Date(s): 03/18/2025 03/18/2025
Heated Purge: (Y/N) N Calibration Time(s): 11:44 14:09
GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF100 = VN085996.D		RRF050 = VN085997.D		RRF020 = VN085998.D			
		RRF010 = VN085999.D		RRF005 = VN086000.D		RRF001 = VN086001.D			
COMPOUND	RRF100	RRF050	RRF020	RRF010	RRF005	RRF001	RRF	% RSD	
t-1,3-Dichloropropene	0.646	0.539	0.504	0.506	0.501	0.446	0.524	12.8	
cis-1,3-Dichloropropene	0.659	0.568	0.533	0.555	0.537	0.464	0.553	11.5	
1,1,2-Trichloroethane	0.382	0.327	0.316	0.330	0.338	0.335	0.338	6.7	
2-Hexanone	0.334	0.284	0.269	0.270	0.261	0.202	0.270	15.7	
Dibromochloromethane	0.503	0.432	0.408	0.422	0.436	0.416	0.436	7.8	
1,2-Dibromoethane	0.400	0.351	0.326	0.343	0.337	0.325	0.347	8	
Tetrachloroethene	0.407	0.371	0.370	0.390	0.421	0.433	0.399	6.6	
Chlorobenzene	1.229	1.085	1.070	1.116	1.156	1.208	1.144	5.7	
Ethyl Benzene	2.209	1.918	1.744	1.769	1.755	1.586	1.830	11.7	
m/p-Xylenes	0.867	0.756	0.700	0.690	0.660	0.643	0.719	11.4	
o-Xylene	0.831	0.713	0.655	0.645	0.585	0.532	0.660	15.8	
Styrene	1.431	1.219	1.115	1.044	1.032	0.924	1.128	15.8	
Bromoform	0.392	0.345	0.329	0.342	0.355	0.371	0.356	6.4	
Isopropylbenzene	3.863	3.599	3.251	3.470	3.264	3.039	3.414	8.6	
1,1,2,2-Tetrachloroethane	1.076	1.052	1.041	1.191	1.261	1.311	1.155	10	
1,3-Dichlorobenzene	1.842	1.685	1.614	1.749	1.718	1.688	1.716	4.5	
1,4-Dichlorobenzene	1.825	1.667	1.621	1.756	1.907	2.043	1.803	8.7	
1,2-Dichlorobenzene	1.716	1.598	1.522	1.673	1.724	2.004	1.706	9.7	
1,2-Dibromo-3-Chloropropane	0.240	0.221	0.216	0.266	0.250	0.192	0.231	11.5	
1,2,4-Trichlorobenzene	0.952	0.850	0.779	0.815	0.851	0.899	0.858	7.1	
1,2,3-Trichlorobenzene	0.900	0.829	0.772	0.794	0.791	0.798	0.814	5.7	
1,2-Dichloroethane-d4	0.632	0.665	0.669	0.721	0.737		0.685	6.3	
Dibromofluoromethane	0.333	0.334	0.347	0.362	0.368		0.349	4.5	
Toluene-d8	1.309	1.266	1.253	1.289	1.217		1.267	2.8	
4-Bromofluorobenzene	0.514	0.466	0.447	0.440	0.391		0.452	9.8	

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q1622 SAS No.: Q1622 SDG No.: Q1622

Instrument ID: MSVOA_N Calibration Date/Time: 03/24/2025 12:03

Lab File ID: VN086065.D Init. Calib. Date(s): 03/18/2025 03/18/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:44 14:09

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.662	0.632		-4.53	20
Chloromethane	0.581	0.530	0.1	-8.78	20
Vinyl Chloride	0.594	0.574		-3.37	20
Bromomethane	0.404	0.365		-9.65	20
Chloroethane	0.375	0.357		-4.8	20
Trichlorofluoromethane	1.067	1.041		-2.44	20
1,1,2-Trichlorotrifluoroethane	0.567	0.571		0.7	20
1,1-Dichloroethene	0.512	0.518		1.17	20
Acetone	0.209	0.194		-7.18	20
Carbon Disulfide	1.685	1.519		-9.85	20
Methyl tert-butyl Ether	1.679	1.722		2.56	20
Methyl Acetate	0.518	0.551		6.37	20
Methylene Chloride	0.612	0.635		3.76	20
trans-1,2-Dichloroethene	0.560	0.579		3.39	20
1,1-Dichloroethane	1.001	1.035	0.1	3.4	20
Cyclohexane	0.853	0.807		-5.39	20
2-Butanone	0.302	0.303		0.33	20
Carbon Tetrachloride	0.604	0.650		7.62	20
cis-1,2-Dichloroethene	0.636	0.674		5.97	20
Bromochloromethane	0.424	0.405		-4.48	20
Chloroform	1.107	1.164		5.15	20
1,1,1-Trichloroethane	1.048	1.089		3.91	20
Methylcyclohexane	0.456	0.445		-2.41	20
Benzene	1.443	1.563		8.32	20
1,2-Dichloroethane	0.503	0.524		4.18	20
Trichloroethene	0.374	0.359		-4.01	20
1,2-Dichloropropane	0.328	0.360		9.76	20
Bromodichloromethane	0.539	0.577		7.05	20
4-Methyl-2-Pentanone	0.371	0.426		14.82	20
Toluene	0.885	1.004		13.45	20
t-1,3-Dichloropropene	0.524	0.553		5.53	20
cis-1,3-Dichloropropene	0.553	0.584		5.61	20
1,1,2-Trichloroethane	0.338	0.376		11.24	20
2-Hexanone	0.270	0.315		16.67	20
Dibromochloromethane	0.436	0.493		13.07	20
1,2-Dibromoethane	0.347	0.376		8.36	20
Tetrachloroethene	0.399	0.398		-0.25	20
Chlorobenzene	1.144	1.142	0.3	-0.17	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q1622 SAS No.: Q1622 SDG No.: Q1622

Instrument ID: MSVOA_N Calibration Date/Time: 03/24/2025 12:03

Lab File ID: VN086065.D Init. Calib. Date(s): 03/18/2025 03/18/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:44 14:09

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.830	1.972		7.76	20
m/p-Xylenes	0.719	0.809		12.52	20
o-Xylene	0.660	0.752		13.94	20
Styrene	1.128	1.324		17.38	20
Bromoform	0.356	0.372	0.1	4.49	20
Isopropylbenzene	3.414	3.371		-1.26	20
1,1,2,2-Tetrachloroethane	1.155	1.071	0.3	-7.27	20
1,3-Dichlorobenzene	1.716	1.699		-0.99	20
1,4-Dichlorobenzene	1.803	1.675		-7.1	20
1,2-Dichlorobenzene	1.706	1.650		-3.28	20
1,2-Dibromo-3-Chloropropane	0.231	0.212		-8.23	20
1,2,4-Trichlorobenzene	0.858	0.759		-11.54	20
1,2,3-Trichlorobenzene	0.814	0.761		-6.51	20
1,2-Dichloroethane-d4	0.685	0.707		3.21	20
Dibromofluoromethane	0.349	0.367		5.16	20
Toluene-d8	1.267	1.355		6.95	20
4-Bromofluorobenzene	0.452	0.494		9.29	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.